



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 9/24/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 16:40
In Date: 09/23/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 09/24/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB105288

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K4664-10	OR-03-092319	20	1.15	9.59	9.11	94.3	
K4664-11	OR-03-092319-1	21	1.12	9.85	9.36	94.4	
K4664-12	OR-03-092319-2	22	1.13	9.82	9.35	94.6	
K4668-12	SU-03-092319	23	1.15	9.61	8.98	92.6	
K4668-13	SU-03-092319-1	24	1.12	9.66	8.89	91.0	
K4668-14	SU-03-092319-2	25	1.12	9.93	9.47	94.8	
K5008-01	SB-1	1	1.12	9.64	8.76	89.7	
K5008-02	SB-2	2	1.12	9.66	8.74	89.2	
K5008-03	SB-3	3	1.12	9.65	8.69	88.7	
K5008-04	SB-4	4	1.12	9.78	8.84	89.1	
K5008-05	SB-5	5	1.14	9.58	8.55	87.8	
K5008-06	SB-6	6	1.12	9.6	8.83	90.9	
K5008-07	SB-7	7	1.16	9.51	8.38	86.5	
K5008-08	SB-8	8	1.14	9.84	8.54	85.1	
K5008-09	SB-9	9	1.14	9.68	8.81	89.8	
K5009-01	HA-06	10	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100% SOLIDS
K5009-02	HA-07	11	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100% SOLIDS
K5009-03	HA-08	12	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100% SOLIDS
K5009-04	HA-09	13	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100% SOLIDS
K5009-05	HA-10	14	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100% SOLIDS
K5009-06	HA-11	15	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K5009-07	HA-12	16	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K5009-08	HA-23	17	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K5010-03	BSP-2	18	1.14	9.66	9.06	93.0	
K5010-04	BSP-2	19	1.14	6.19	5.78	91.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$