


LB105289

ISM02.4

INSTRUMENT ID : CU1



Linear

A= 0.0000e+000

B= 2.4105e-004

C= 1.3472e-001

Rho= 0.9998996

Accept = Accepted

slope
y intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.075	-0.075	-868	0.000	-868				
0.05	0.050									
0.20	0.200	0.256	0.056	505	0.0 %	505				
2.50	2.500	2.476	-0.024	9712	0.0 %	9712				
5.00	5.000	5.073	0.073	20486	0.0 %	20486				
7.50	7.500	7.510	0.010	30599	0.0 %	30599				
10.0	10.000	9.959	-0.041	40757	0.0 %	40757				

 16D
 1
 28
 1
 00

NS 09-23-19

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	-868	-		0 ISM02.4	PPB	9/23/2019 15:28	Std
	0.2 S0.2	505	-		0.2 ISM02.4	PPB	9/23/2019 15:30	Std
	2.5 S2.5	9712	-		2.5 ISM02.4	PPB	9/23/2019 15:33	Std
	5 S5.0	20486	-		5 ISM02.4	PPB	9/23/2019 15:35	Std
	7.5 S7.5	30599	-		7.5 ISM02.4	PPB	9/23/2019 15:37	Std
	10 S10.0	40757	-		10 ISM02.4	PPB	9/23/2019 15:43	Std
ICV	ICV	16633	4.144 -		ISM02.4	PPB	9/23/2019 15:51	SMPL
ICB	ICB	91	0.1567 -		ISM02.4	PPB	9/23/2019 15:53	SMPL
CCV094	CCV094	21098	5.2203 -		ISM02.4	PPB	9/23/2019 15:55	SMPL
CCB094	CCB094	49	0.1465 -		ISM02.4	PPB	9/23/2019 15:57	SMPL
PB123259BL	PBW259	41	0.1446 -		ISM02.4	PPB	9/23/2019 16:10	SMPL
K4735-01	MH0FD6	102611	24.8687 -		ISM02.4	PPB	9/23/2019 16:12	SMPL
K4735-02	MH0FD7	444982	HIGH -		ISM02.4	PPB	9/23/2019 16:14	
K4735-03	MH0FE1	140812	34.0769 -		ISM02.4	PPB	9/23/2019 16:16	SMPL
K4735-04	MH0FE7	4353	1.184 -		ISM02.4	PPB	9/23/2019 16:18	SMPL
K4735-05	MH0FF1	11502	2.9072 -		ISM02.4	PPB	9/23/2019 16:21	SMPL
K4735-06	MH0FF2	13534	3.397 -		ISM02.4	PPB	9/23/2019 16:24	SMPL
K4735-07	MH0FF7	845	0.3384 -		ISM02.4	PPB	9/23/2019 16:26	SMPL
K4735-08	MH0FF8	15885	3.9637 -		ISM02.4	PPB	9/23/2019 16:29	SMPL
K4735-09	MH0FG0	80195	19.4654 -		ISM02.4	PPB	9/23/2019 16:31	SMPL
K4735-10	MH0FG6	649	0.2912 -		ISM02.4	PPB	9/23/2019 16:33	SMPL
K4735-11	MH0FG8	19042	4.7247 -		ISM02.4	PPB	9/23/2019 16:36	SMPL
CCV095	CCV095	20704	5.1253 -		ISM02.4	PPB	9/23/2019 16:38	SMPL
CCB095	CCB095	144	0.1694 -		ISM02.4	PPB	9/23/2019 16:40	SMPL
K4735-12	MH0FG9	13995	3.5082 -		ISM02.4	PPB	9/23/2019 16:42	SMPL
K4735-13	MH0FG9D	13781	3.4566 -		ISM02.4	PPB	9/23/2019 16:45	SMPL
K4735-14	MH0FG9S	25051	6.1732 -		ISM02.4	PPB	9/23/2019 16:47	SMPL
K4735-01DLX5	MH0FD6	20011	4.9583 -		ISM02.4	PPB	9/23/2019 16:49	SMPL
K4735-02DLX15	MH0FD7	42184	10.303 -		ISM02.4	PPB	9/23/2019 16:51	SMPL
K4735-03DLX5	MH0FE1	27739	6.8211 -		ISM02.4	PPB	9/23/2019 16:54	SMPL
K4735-09DLX5	MH0FG0	15775	3.9372 -		ISM02.4	PPB	9/23/2019 16:56	SMPL
K4735-02DL2X30	MH0FD7	20774	5.1422 -		ISM02.4	PPB	9/23/2019 16:58	SMPL
PB123259TB	LEB259	141	0.1687 -		ISM02.4	PPB	9/23/2019 17:03	SMPL
CCV096	CCV096	19910	4.9339 -		ISM02.4	PPB	9/23/2019 17:06	SMPL
CCB096	CCB096	-12	0.1318 -		ISM02.4	PPB	9/23/2019 17:08	SMPL