



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB105291

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/23/2019	17:00
2	CAL2	1	Water	NA	NA	20.1	7.01	09/23/2019	17:01
3	CAL3	1	Water	NA	NA	20.1	10.03	09/23/2019	17:02
4	ICV	1	Water	NA	NA	20.2	7.00	09/23/2019	17:05
5	CCV1	1	Water	NA	NA	20.1	2.02	09/23/2019	17:06
6	K4664-13	1	Solid	20.02	20	23.9	7.45	09/23/2019	17:10
7	K4664-13DUP	1	Solid	20.03	20	23.8	7.46	09/23/2019	17:11
8	K4665-06	1	Solid	20.04	20	21.7	8.03	09/23/2019	17:12
9	K4668-15	1	Solid	20.01	20	24.2	7.85	09/23/2019	17:15
10	K4997-11	1	Solid	20.03	20	22.1	7.82	09/23/2019	17:16
11	K4997-13	1	Solid	20.05	20	21.7	7.50	09/23/2019	17:18
12	K4997-15	1	Solid	20.03	20	21.8	7.28	09/23/2019	17:20
13	K4997-17	1	Solid	20.02	20	21.2	7.07	09/23/2019	17:22
14	K4997-19	1	Solid	20.01	20	21.3	6.93	09/23/2019	17:25
15	K5010-01	1	Solid	20.03	20	23.0	5.74	09/23/2019	17:30
16	CCV2	1	Water	NA	NA	20.1	12.01	09/23/2019	17:31

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K4668 WorkList ID : 131238 Department : Wet-Chemistry Date : 09-23-2019 12:09:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/23/2019	Solid	K4664-13-0	Corrosivity	Cool 4 deg C	PSEG05		OR-03-092319	09/23/2019	9045D
09/19/2019	Solid	K4665-06-0	Corrosivity	Cool 4 deg C	PSEG05	N22	EO-01-091919	09/19/2019	9045D
09/23/2019	Solid	K4668-15-0	Corrosivity	Cool 4 deg C	PSEG05		SU-03-092319	09/23/2019	9045D
09/19/2019	Solid	K4997-11-0	Corrosivity	Cool 4 deg C	SEVE02	N21	WC-08-091919	09/19/2019	9045D
09/19/2019	Solid	K4997-13-0	Corrosivity	Cool 4 deg C	SEVE02	N21	WC-09-091919	09/19/2019	9045D
09/19/2019	Solid	K4997-15-0	Corrosivity	Cool 4 deg C	SEVE02	N21	WC-10-091919	09/19/2019	9045D
09/19/2019	Solid	K4997-17-0	Corrosivity	Cool 4 deg C	SEVE02	N21	WC-11-091919	09/19/2019	9045D
09/19/2019	Solid	K4997-19-0	Corrosivity	Cool 4 deg C	SEVE02	N21	WC-12-091919	09/19/2019	9045D
09/23/2019	Solid	K5010-01-0	Corrosivity	Cool 4 deg C	AREC01	N61	BSP-1	09/23/2019	9045D

Date/Time 09/23/19 18:00
 Received by: Ja
 Relinquished by: CD

Date/Time 09/23/19 18:00
 Received by: CD
 Relinquished by: Ja