



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB105298

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP78296
Turbidity Calibration std, 0NTU	WP78290
Turbidity Calibration std, 20NTU	WP78292
Turbidity Calibration std, 40NTU	WP78291
Turbidity Calibration std, 1NTU	WP78295
Turbidity Calibration std, 10NTU	WP78293
Turbidity Calibration std, 5NTU	WP78294
10 NTU Standard 500 ml	W2489

Intercept: -0.0026

Slope: 0.9391

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.175	09/24/2019	09:45
2	CAL2	1	1	1.050	09/24/2019	09:52
3	CAL3	5	1	4.610	09/24/2019	09:57
4	CAL4	10	1	9.220	09/24/2019	10:01
5	CAL5	20	1	18.600	09/24/2019	10:05
6	CAL6	40	1	37.700	09/24/2019	10:09



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB105298

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.600	10.225	09/24/2019	10:14
2	ICB		1	0.179	0.193	09/24/2019	10:19
3	CCV1	10	1	9.160	9.757	09/24/2019	10:24
4	CCB1		1	0.177	0.191	09/24/2019	10:29
5	LB105298BL		1	0.172	0.186	09/24/2019	10:34
6	K4843-37		1	18.500	19.702	09/24/2019	10:39
7	CCV2	10	1	9.180	9.778	09/24/2019	10:43
8	CCB2		1	0.180	0.194	09/24/2019	10:48