



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 9/25/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 17:11
In Date: 09/24/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 09/25/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB105307

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K4657-01	BU-02-092419	1	1.13	9.86	7.67	74.9	
K4657-02	BU-02-092419	2	1.14	9.78	7.29	71.2	
K4657-04	BU-03-092419	3	1.12	9.57	8.9	92.1	
K4657-05	BU-03-092419	4	1.17	9.6	8.31	84.7	
K4663-07	OK-02-092419	5	1.12	9.58	9.15	94.9	
K4663-08	OK-02-092419	6	1.14	9.53	8.85	91.9	
K5012-02	IDW-GAC1-190923	7	1.14	9.76	5.33	48.6	
K5014-01	BASEMENT-SLAB	8	1.00	2.00	2.00	100.0	CONCRETE SAMPLE 100 % SOLIDS
K5014-02	WALL-1	9	1.00	2.00	2.00	100.0	CONCRETE SAMPLE 100 % SOLIDS
K5014-03	WALL-2	10	1.00	2.00	2.00	100.0	CONCRETE SAMPLE 100 % SOLIDS
K5014-04	SHED-SLAB	11	1.00	2.00	2.00	100.0	CONCRETE SAMPLE 100 % SOLIDS
K5018-01	HP-1	12	1.14	9.73	7.42	73.1	
K5018-02	HP-2	13	1.15	9.87	8.67	86.2	
K5018-03	HP-3	14	1.13	9.7	8.12	81.6	
K5020-01	RT-3860	15	1.15	9.86	9.37	94.4	
K5020-03	RT-3465	16	1.14	9.95	9.29	92.5	
K5020-05	RT-5124	17	1.14	9.6	7.86	79.4	
K5020-07	RBR-200050	18	1.16	9.7	8.56	86.7	
K5021-01	RT-1488-RBR200031-RBR200036	19	1.15	9.98	8.42	82.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$