



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB105308

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/24/2019	17:10
2	CAL2	1	Water	NA	NA	20.1	6.99	09/24/2019	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	09/24/2019	17:12
4	ICV	1	Water	NA	NA	20.1	7.01	09/24/2019	17:15
5	CCV1	1	Water	NA	NA	20.2	2.01	09/24/2019	17:16
6	K4657-03	1	Solid	20.02	20	23.4	7.49	09/24/2019	17:20
7	K4657-03DUP	1	Solid	20.03	20	23.3	7.50	09/24/2019	17:21
8	K4657-06	1	Solid	20.04	20	23.4	8.45	09/24/2019	17:22
9	K4663-09	1	Solid	20.03	20	21.9	8.98	09/24/2019	17:25
10	K5012-01	1	Solid	20.02	20	23.4	7.96	09/24/2019	17:26
11	CCV2	1	Water	NA	NA	20.2	12.02	09/24/2019	17:30

YB 105208

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K4983 WorkList ID : 131277 Department : Wet-Chemistry Date : 09-24-2019 16:28:48

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/24/2019	Solid	K4657-03	Corrosivity	Cool 4 deg C	PSEG05	N13	BU-02-092419	09/24/2019	9045D
09/24/2019	Solid	K4657-06	Corrosivity	Cool 4 deg C	PSEG05	N13	BU-03-092419	09/24/2019	9045D
09/24/2019	Solid	K4663-09	Corrosivity	Cool 4 deg C	PSEG05		OK-02-092419	09/24/2019	9045D
09/23/2019	Solid	K5012-01	Corrosivity	Cool 4 deg C	SHAW22		IDW-GAC1-190923	09/23/2019	9045D

Date/Time 09.24.19 16:30
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 09.24.19 18:00
Received by: [Signature]
Relinquished by: [Signature]