



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB105316
Analysis Date: 09/25/2019
BalanceID: WC SC-3
OvenID: EXT OVEN-3

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 09/25/2019
Extraction IN Time: 12:10
Extraction OUT Time: 12:40

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB105316BL	LB105316BL	WATER	1.3	1000	100	2.4016	2.4016	3.02	2.4016	2.4016	0.0000	0
2	LB105316BS	LB105316BS	WATER	1.3	1000	100	2.8634	2.8634	3.01	2.8823	2.8823	0.0189	18.9
3	LB105316BSD	LB105316BSD	WATER	1.3	1000	100	2.8103	2.8103	3.03	2.8293	2.8293	0.0190	19
4	K4843-18	RR-TPH1L-WP	WATER	1.3	880	100	3.0063	3.0063	3.03	3.0935	3.0935	0.0872	99.09
5	K5023-01	GRAB	WATER	1.6	1000	100	2.9948	2.9948	3.05	3.0359	3.0359	0.0411	41.1



QC Batch# LB105316 Test: TPH Analysis Date: 09/25/2019

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2591
pH Paper 0-14	M4240
Sodium Sulfate	EP1987
1:1 HCL	WP74116
Silica Gel	W2546
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP74117
LCSWD	10.00 ML	WP74118
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>14:11</u>
1.0000 gram Balance:	<u>1.0004</u>	In Time:	<u>13:10</u>		
Bal Check Time:	<u>12:30</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>14:40</u>
		Out Time:	<u>14:10</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>15:46</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>15:10</u>		
Bal Check Time:	<u>16:35</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>16:10</u>
		Out Time:	<u>15:45</u>		

LB 105316

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH K5023 WorkList ID : 131305 Department : Wet-Chemistry Date : 09-25-2019 11:58:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/02/2019	Water	K5023-01 ✓	TPH	Conc H2SO4 to	ARAM01	N22	GRAB	09/25/2019	1664A

Date/Time 09/25/19 12:00
Received by: JP
Relinquished by: CO

Date/Time 09/25/19 13:20
Received by: CO
Relinquished by: JP