



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : AMANDEEP

Run Number: LB105324

Slope : 98.3

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/25/2019	17:05
2	CAL2	1	Water	NA	NA	20.1	7.01	09/25/2019	17:06
3	CAL3	1	Water	NA	NA	20.1	10.03	09/25/2019	17:08
4	ICV	1	Water	NA	NA	20.1	7.00	09/25/2019	17:10
5	CCV1	1	Water	NA	NA	20.1	2.02	09/25/2019	17:11
6	K5037-01	1	Water	NA	NA	22.3	8.64	09/25/2019	17:12
7	K5037-01DUP	1	Water	NA	NA	22.4	8.65	09/25/2019	17:13
8	K5039-10	1	Water	NA	NA	21.4	5.49	09/25/2019	17:15
9	K5039-11	1	Water	NA	NA	21.1	5.47	09/25/2019	17:16
10	CCV2	1	Water	NA	NA	20.1	12.02	09/25/2019	17:20

LB 105324

WORKLIST(Hardcopy Internal Chain)

Date: 09-25-2019 16:19:49

Department: Wet-Chemistry

WorkList ID: 131310

WorkList Name: PH K5037 WATER

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/25/2019	Water	K5037-01 ✓	pH	Cool 4 deg C	PSEG01	N13	40241	09/25/2019	9040C
09/25/2019	Water	K5039-10 ✓	pH	Cool 4 deg C	PSEG01	N13	30451	09/25/2019	9040C
09/25/2019	Water	K5039-11 ✓	pH	Cool 4 deg C	PSEG01	N13	30452	09/25/2019	9040C

Date/Time 09/25/19 16:20

Received by: JA

Relinquished by: DK

Date/Time 09/25/19 18:20

Received by: DK

Relinquished by: JA