



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105342

Slope : 98.3

BalanceID: WC SC-3

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/26/2019	14:15
2	CAL2	1	Water	NA	NA	20.1	7.01	09/26/2019	14:16
3	CAL3	1	Water	NA	NA	20.1	10.04	09/26/2019	14:17
4	ICV	1	Water	NA	NA	20.1	7.00	09/26/2019	14:20
5	CCV1	1	Water	NA	NA	20.1	2.02	09/26/2019	14:22
6	K5060-01	1	Solid	20.02	20	21.4	6.31	09/26/2019	14:25
7	K5060-01DUP	1	Solid	20.03	20	21.5	6.32	09/26/2019	14:26
8	CCV2	1	Water	NA	NA	20.2	12.02	09/26/2019	14:30

VB 105342

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5060 WorkList ID : 131347 Department : Wet-Chemistry Date : 09-26-2019 13:56:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/26/2019	Solid	K5060-01	pH	Cool 4 deg C	TULL02	N53	FK-PS-01-060	09/26/2019	9045D

Date/Time 09/26/19 14:00
Received by: JP
Relinquished by: JP

Date/Time 09/26/19 15:00
Received by: JP
Relinquished by: JP