



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB105345

Slope : 98.4

BalanceID: WC SC-3

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/26/2019	17:10
2	CAL2	1	Water	NA	NA	20.1	7.01	09/26/2019	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	09/26/2019	17:12
4	ICV	1	Water	NA	NA	20.1	6.99	09/26/2019	17:15
5	CCV1	1	Water	NA	NA	20.1	2.02	09/26/2019	17:16
6	K4668-19	1	Solid	20.02	20	22.6	8.50	09/26/2019	17:20
7	K4668-19DUP	1	Solid	20.03	20	22.7	8.51	09/26/2019	17:21
8	K5064-01	1	Solid	20.04	20	23.0	8.77	09/26/2019	17:22
9	CCV2	1	Water	NA	NA	20.1	12.02	09/26/2019	17:25

WORKLIST(Hardcopy Internal Chain)

18765345

WorkList Name : CORROSIONITY K4669 WorkList ID : 131352 Department : Wet-Chemistry Date : 09-26-2019 16:37:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/26/2019	Solid	K4668-19 -P	Corrosivity	Cool 4 deg C	PSEG05		SU-04-092619	09/26/2019	9045D
09/26/2019	Solid	K5064-01 -P	Corrosivity	Cool 4 deg C	PSEG01	N13	DRUM-COMP	09/26/2019	9045D

Date/Time 09/26/19 16:40
 Received by: JF
 Relinquished by: JF

Date/Time 09/26/19 17:40
 Received by: CD
 Relinquished by: JF