



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 9/27/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 17:35
In Date: 09/26/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:40
Out Date: 09/27/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB105346

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K5059-01	MDBAD5	1	1.00	2.00	2.00	100.0	P.E.SAMPLE
K5059-02	MDBAD8	2	1.15	9.88	6.71	63.7	
K5059-03	MDBAE0	3	1.12	9.76	8.48	85.2	
K5059-04	MDBAE1	4	1.16	9.81	7.22	70.1	
K5059-05	MDBAE2	5	1.18	9.86	7.06	67.7	
K5059-06	MDBAE3	6	1.15	9.94	8.2	80.2	
K5059-07	MDBAE5	7	1.15	9.96	7.17	68.3	
K5059-08	MDBAE6	8	1.15	9.71	7.94	79.3	
K5059-09	MDBAE9	9	1.11	9.72	8.29	83.4	
K5059-10	MDBAE9D	10	1.11	9.72	8.29	83.4	
K5059-11	MDBAE9S	11	1.11	9.72	8.29	83.4	
K5059-12	MDBAF0	12	1.16	9.58	8.3	84.8	
K5059-13	MDBAF1	13	1.14	9.87	9.00	90.0	
K5059-14	MDBAF2	14	1.16	9.77	7.93	78.6	
K5059-15	MDBAF6	15	1.19	9.82	8.19	81.1	
K5059-16	MDBAG5	16	1.13	9.65	6.68	65.1	
K5059-17	MDBAG6	17	1.16	9.97	7.4	70.8	
K5059-18	MDBAG7	18	1.16	9.94	8.65	85.3	
K5059-19	MDBAH0	19	1.14	9.74	6.11	57.8	
K5059-20	MDBAH1	20	1.16	9.9	7.39	71.3	
K5059-21	MDBAH2	21	1.17	9.72	6.82	66.1	
K5059-22	MDBAH4	22	1.15	9.52	7.09	71.0	
K5059-23	MDBAH7	23	1.17	9.83	6.41	60.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$