


LB105351

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.4468e-004

C= -5.8985e-002

Rho= 0.9999022

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.045	-0.045	56	0.000	56				
0.2	0.200	0.203	0.003	1072	0.0 %	1072				
2.5	2.500	2.499	-0.001	10453	0.0 %	10453				
5.0	5.000	5.108	0.108	21119	0.0 %	21119				
7.5	7.500	7.462	-0.038	30738	0.0 %	30738				
10.0	10.000	9.973	-0.027	41001	0.0 %	41001				

LB105351 INSTRUMENT ID : CV1

Method: 7470A	Operator: Admin	Date of Analysis: 27 Sep 2019 11:02:41							
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	56	-	0.0000	7470A	PPB	27 Sep 2019 11:08:49	S	Std
0.2 - 1	50-2	1072	-	0.2000	7470A	PPB	27 Sep 2019 11:10:58	S	Std
2.5 - 1	52-5	10453	-	2.5000	7470A	PPB	27 Sep 2019 11:13:07	S	Std
5.0 - 1	55-0	21119	-	5.0000	7470A	PPB	27 Sep 2019 11:15:15	S	Std
7.5 - 1	57-5	30738	-	7.5000	7470A	PPB	27 Sep 2019 11:17:23	S	Std
10.0 - 1	510-0	41001	-	10.0000	7470A	PPB	27 Sep 2019 11:19:31	S	Std
ICV31 - 1	ICV31	16287	3.9261	-	7470A	PPB	27 Sep 2019 11:26:37	U	SMPL
ICB31 - 1	ICB31	37	-0.0499	-	7470A	PPB	27 Sep 2019 11:28:44	U	SMPL
CCV16 - 1	CCV16	21199	5.1279	-	7470A	PPB	27 Sep 2019 11:30:51	U	SMPL
CCB16 - 1	CCB16	14	-0.0556	-	7470A	PPB	27 Sep 2019 11:32:57	U	SMPL
CRA - 1	CRA	1073	0.2036	-	7470A	PPB	27 Sep 2019 11:35:04	U	SMPL
HighStd - 1	HighStd	40852	9.9366	-	7470A	PPB	27 Sep 2019 11:37:11	U	SMPL
ChkStd - 1	ChkStd	28396	6.8889	-	7470A	PPB	27 Sep 2019 11:39:18	U	SMPL
PB123438BL - 1	PBW	-18	-0.0634	-	7470A	PPB	27 Sep 2019 11:41:25	U	SMPL
PB123438BS - 1	LCSW	16340	3.9391	-	7470A	PPB	27 Sep 2019 11:43:33	U	SMPL
K4657-03 - 1	BU-02-092419	17	-0.0548	-	7470A	PPB	27 Sep 2019 11:45:40	U	SMPL
K4657-03DUP - 1	BU-02-092419DUP	41	-0.0490	-	7470A	PPB	27 Sep 2019 11:47:48	U	SMPL
K4657-03MSD - 1	BU-02-092419MSD	16279	3.9241	-	7470A	PPB	27 Sep 2019 11:52:05	U	SMPL
K4657-06 - 1	BU-03-092419	0	-0.0590	-	7470A	PPB	27 Sep 2019 11:54:12	U	SMPL
CCV17 - 1	CCV17	21190	5.1257	-	7470A	PPB	27 Sep 2019 11:56:18	U	SMPL
CCB17 - 1	CCB17	4	-0.0580	-	7470A	PPB	27 Sep 2019 11:58:25	U	SMPL
K4663-09 - 1	OK-02-092419	50	-0.0468	-	7470A	PPB	27 Sep 2019 12:00:32	U	SMPL
K5012-01 - 1	IDW-GAC1-190923	63	-0.0436	-	7470A	PPB	27 Sep 2019 12:02:39	U	SMPL
K5020-02 - 1	RT-3860	46	-0.0477	-	7470A	PPB	27 Sep 2019 12:04:46	U	SMPL
K5020-04 - 1	RT-3465	56	-0.0453	-	7470A	PPB	27 Sep 2019 12:06:53	U	SMPL
K5020-06 - 1	RT-5124	47	-0.0475	-	7470A	PPB	27 Sep 2019 12:09:01	U	SMPL
K5020-08 - 1	RBR-200050	17	-0.0548	-	7470A	PPB	27 Sep 2019 12:11:09	U	SMPL
PB123439BL - 1	PBW	24	-0.0531	-	7470A	PPB	27 Sep 2019 12:13:17	U	SMPL
PB123439BS - 1	LCSW	16444	3.9645	-	7470A	PPB	27 Sep 2019 12:15:25	U	SMPL
K5037-02 - 1	30445	28	-0.0521	-	7470A	PPB	27 Sep 2019 12:17:34	U	SMPL
K5039-02 - 1	30476-80	23	-0.0534	-	7470A	PPB	27 Sep 2019 12:19:40	U	SMPL
CCV18 - 1	CCV18	21229	5.1353	-	7470A	PPB	27 Sep 2019 12:21:47	U	SMPL
CCB18 - 1	CCB18	4	-0.0580	-	7470A	PPB	27 Sep 2019 12:23:55	U	SMPL
K5039-12 - 1	40245-247-248	31	-0.0514	-	7470A	PPB	27 Sep 2019 12:26:02	U	SMPL
K5039-12DUP - 1	40245-247-248DUP	42	-0.0487	-	7470A	PPB	27 Sep 2019 12:28:08	U	SMPL
K5039-12MS - 1	40245-247-248MS	16666	4.0188	-	7470A	PPB	27 Sep 2019 12:30:15	U	SMPL
K5039-12MSD - 1	40245-247-248MSD	16309	3.9315	-	7470A	PPB	27 Sep 2019 12:32:22	U	SMPL
K5039-16 - 1	40246	76	-0.0404	-	7470A	PPB	27 Sep 2019 12:34:30	U	SMPL
PB123440BL - 1	PBW	145	-0.0235	-	7470A	PPB	27 Sep 2019 12:36:37	U	SMPL
PB123440BS - 1	LCSW	16200	3.9048	-	7470A	PPB	27 Sep 2019 12:38:46	U	SMPL
K5054-01 - 1	PAT-091019-B	222	-0.0047	-	7470A	PPB	27 Sep 2019 12:40:54	U	SMPL
K5054-01DUP - 1	PAT-091019-BDUP	237	-0.0010	-	7470A	PPB	27 Sep 2019 12:43:03	U	SMPL
K5054-01MS - 1	PAT-091019-BMS	15926	3.8378	-	7470A	PPB	27 Sep 2019 12:45:10	U	SMPL
CCV19 - 1	CCV19	21261	5.1431	-	7470A	PPB	27 Sep 2019 12:47:17	U	SMPL
CCB19 - 1	CCB19	101	-0.0343	-	7470A	PPB	27 Sep 2019 12:49:24	U	SMPL
K5054-01MSD - 1	PAT-091019-BMSD	15684	3.7785	-	7470A	PPB	27 Sep 2019 12:51:31	U	SMPL
K5054-02 - 1	PAT-091019-C	358	0.0286	-	7470A	PPB	27 Sep 2019 12:53:38	U	SMPL
PB123441BL - 1	PBW	32	-0.0512	-	7470A	PPB	27 Sep 2019 12:55:45	U	SMPL
PB123441BS - 1	LCSW	16415	3.9574	-	7470A	PPB	27 Sep 2019 13:03:11	U	SMPL
K5003-01 - 1	TMW-03	174	-0.0164	-	7470A	PPB	27 Sep 2019 13:05:18	U	SMPL
K5003-01DUP - 1	TMW-03DUP	209	-0.0078	-	7470A	PPB	27 Sep 2019 13:07:27	U	SMPL
K5003-01MSD - 1	TMW-03MSD	15648	3.7697	-	7470A	PPB	27 Sep 2019 13:11:44	U	SMPL
PB123438TB - 1	PB123364TB	27	-0.0524	-	7470A	PPB	27 Sep 2019 13:13:53	U	SMPL
PB123439TB - 1	PB123387TB	32	-0.0512	-	7470A	PPB	27 Sep 2019 13:16:01	U	SMPL
CCV20 - 1	CCV20	20698	5.0054	-	7470A	PPB	27 Sep 2019 13:18:08	U	SMPL
CCB20 - 1	CCB20	120	-0.0296	-	7470A	PPB	27 Sep 2019 13:20:16	U	SMPL
PB123440TB - 1	PB123418TB	47	-0.0475	-	7470A	PPB	27 Sep 2019 13:22:24	U	SMPL
K4657-03LX5 - 1		37	-0.0499	-	7470A	PPB	27 Sep 2019 13:24:31	U	SMPL
K4657-03A - 1		16291	3.9271	-	7470A	PPB	27 Sep 2019 13:26:39	U	SMPL
K5039-12LX5 - 1		6	-0.0575	-	7470A	PPB	27 Sep 2019 13:28:46	U	SMPL
K5039-12A - 1		16632	4.0105	-	7470A	PPB	27 Sep 2019 13:30:54	U	SMPL
K5054-01LX5 - 1		23	-0.0534	-	7470A	PPB	27 Sep 2019 13:33:02	U	SMPL
K5054-01A - 1		15708	3.7844	-	7470A	PPB	27 Sep 2019 13:35:11	U	SMPL
K5003-01LX5 - 1		3	-0.0583	-	7470A	PPB	27 Sep 2019 13:37:19	U	SMPL
K4657-03 - 1	BU-02-092419MS	16414	3.9572	-	7470A	PPB	27 Sep 2019 13:41:36	U	SMPL
CCV21 - 1	CCV21	21223	5.1338	-	7470A	PPB	27 Sep 2019 13:43:43	U	SMPL
CCB21 - 1	CCB21	-19	-0.0636	-	7470A	PPB	27 Sep 2019 13:45:51	U	SMPL
K5003-01MS - 1	TMW-03MS	16008	3.8578	-	7470A	PPB	27 Sep 2019 13:47:59	U	SMPL
K5003-01A - 1		16224	3.9107	-	7470A	PPB	27 Sep 2019 13:50:08	U	SMPL
CCV22 - 1	CCV22	21494	5.2001	-	7470A	PPB	27 Sep 2019 13:52:18	U	SMPL
CCB22 - 1	CCB22	7	-0.0573	-	7470A	PPB	27 Sep 2019 13:54:26	U	SMPL