



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105354

Slope : 98.1

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	3.99	09/27/2019	13:10
2	CAL2	1	Water	NA	NA	20.2	7.00	09/27/2019	13:11
3	CAL3	1	Water	NA	NA	20.2	10.03	09/27/2019	13:12
4	ICV	1	Water	NA	NA	20.2	7.01	09/27/2019	13:15
5	CCV1	1	Water	NA	NA	20.1	2.01	09/27/2019	13:16
6	K5090-01	1	Water	NA	NA	21.6	6.02	09/27/2019	13:19
7	K5090-01DUP	1	Water	NA	NA	21.8	6.03	09/27/2019	13:20
8	CCV2	1	Water	NA	NA	20.2	12.01	09/27/2019	13:22

LB103354

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5090 WorkList ID : 131387 Department : Wet-Chemistry Date : 09-27-2019 12:53:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/26/2019	Water	K5090-01-C	pH	Cool 4 deg C	STER01	N31	POOL-WATER	09/26/2019	9040C

Date/Time 09/27/19 12:00
Received by: JR
Relinquished by: CJ

Date/Time 09/27/19 14:00
Received by: CJ
Relinquished by: JR