



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105363

Slope : 98.4

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/27/2019	14:30
2	CAL2	1	Water	NA	NA	20.1	7.01	09/27/2019	14:31
3	CAL3	1	Water	NA	NA	20.1	10.03	09/27/2019	14:32
4	ICV	1	Water	NA	NA	20.1	7.00	09/27/2019	14:35
5	CCV1	1	Water	NA	NA	20.2	2.02	09/27/2019	14:36
6	K5088-01	1	Water	NA	NA	25.1	9.43	09/27/2019	14:37
7	K5088-01DUP	1	Water	NA	NA	25.2	9.44	09/27/2019	14:38
8	CCV2	1	Water	NA	NA	20.1	12.02	09/27/2019	14:40

20105363

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5088 WorkList ID : 131393 Department : Wet-Chemistry Date : 09-27-2019 14:21:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/27/2019	Water	K5088-01 -PT	pH	Cool 4 deg C	PSEG01	N13	SP-NATURAL-GAS-HEATER	09/27/2019	9040C

Date/Time 09/27/19 14:25
Received by: Jo
Relinquished by: MP

Date/Time 09/27/19 15:00
Received by: MP
Relinquished by: Jo