



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB105365

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	09/27/2019	14:55
2	K5037-01		1	9	100.00	>212.0	>212.0	09/27/2019	15:20
3	K5037-01DUP		1	9	100.00	>212.0	>212.0	09/27/2019	15:45
4	K5054-01		1	7	100.00	>212.0	>212.0	09/27/2019	16:10
5	K5054-02		1	8	100.00	>212.0	>212.0	09/27/2019	16:35
6	K5088-01		1	9	100.00	>212.0	>212.0	09/27/2019	16:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB105365

WorkList Name : FP92719 WorkList ID : 131394 Department : Wet-Chemistry Date : 09-27-2019 14:24:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/02/2019	Water	K5037-01 <i>P</i>	Flash Point	Cool 4 deg C	PSEG01	N13	40241	09/25/2019	1010A
10/01/2019	Water	K5054-01 <i>K</i>	Flash Point	Cool 4 deg C	PSEG01	N13	PAT-091019-B	09/26/2019	1010A
10/01/2019	Water	K5054-02 <i>E</i>	Flash Point	Cool 4 deg C	PSEG01	N13	PAT-091019-C	09/26/2019	1010A
09/30/2019	Water	K5088-01 <i>A</i>	Flash Point	Cool 4 deg C	PSEG01	N13	SP-NATURAL-GAS-HEATER	09/27/2019	1010A

Date/Time 09/27/19 14:25
 Received by: *K5*
 Relinquished by: *CP*

Date/Time 09/27/19 17:00
 Received by: *CP*
 Relinquished by: *K5*