



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB105366

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	09/27/2019	17:20
2	K5039-19		1	8	60.00	140.0	140.0	09/27/2019	17:40
3	K5039-19DUP		1	8	60.00	140.0	140.0	09/27/2019	18:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB105366

WorkList Name : FP92719-1 WorkList ID : 131395 Department : Wet-Chemistry Date : 09-27-2019 14:25:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/02/2019	Solid	K5039-19	Flash Point	Cool 4 deg C	PSEG01	N13	30430	09/25/2019	1010A

Date/Time 09/27/19 14:30
Received by: KS
Relinquished by: CP

Date/Time 09/27/19 18:10
Received by: KS
Relinquished by: KS