



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB105392

Slope : 99.2

pH Meter ID : WC PH METER

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/30/2019	15:05
2	CAL2	1	Water	NA	NA	20.1	7.00	09/30/2019	15:06
3	CAL3	1	Water	NA	NA	20.1	10.03	09/30/2019	15:07
4	ICV	1	Water	NA	NA	20.1	7.01	09/30/2019	15:09
5	CCV1	1	Water	NA	NA	20.1	2.01	09/30/2019	17:10
6	K5095-04	1	Water	NA	NA	19.2	7.29	09/30/2019	17:11
7	K5095-04DUP	1	Water	NA	NA	19.1	7.30	09/30/2019	17:12
8	K5095-05	1	Water	NA	NA	19.9	7.39	09/30/2019	17:13
9	CCV2	1	Water	NA	NA	20.1	12.02	09/30/2019	17:15

LB 105392

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K5095 W WorkList ID : 131455 Department : Wet-Chemistry Date : 09-30-2019 14:54:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/27/2019	Water	K5095-04 -D	Corrosivity	Cool 4 deg C	SEAC01	O41	COMPOSITE-4	09/27/2019	9040C
09/27/2019	Water	K5095-05 -D	Corrosivity	Cool 4 deg C	SEAC01	O41	COMPOSITE-5	09/27/2019	9040C

Date/Time 09/30/19 13:00
 Received by: Jz
 Relinquished by: cd

Date/Time 09/30/19 16:100
 Received by: cd
 Relinquished by: Jz