



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB105393

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/30/2019	15:05
2	CAL2	1	Water	NA	NA	20.1	7.00	09/30/2019	15:06
3	CAL3	1	Water	NA	NA	20.1	10.03	09/30/2019	15:07
4	ICV	1	Water	NA	NA	20.1	7.01	09/30/2019	15:09
5	CCV1	1	Water	NA	NA	20.1	2.01	09/30/2019	17:25
6	K4659-03	1	Solid	20.02	20	22.2	9.67	09/30/2019	17:26
7	K4659-03DUP	1	Solid	20.03	20	22.1	9.68	09/30/2019	17:27
8	K4659-06	1	Solid	20.04	20	21.2	9.45	09/30/2019	17:30
9	K4663-12	1	Solid	20.02	20	22.3	9.02	09/30/2019	17:32
10	K4946-16	1	Solid	20.03	20	21.2	8.52	09/30/2019	17:33
11	K4946-21	1	Solid	20.02	20	21.2	7.12	09/30/2019	17:35
12	K5095-01	1	Solid	20.02	20	21.2	8.50	09/30/2019	17:36
13	K5095-02	1	Solid	20.04	20	21.7	5.37	09/30/2019	17:39
14	K5095-03	1	Solid	20.02	20	21.4	6.45	09/30/2019	17:40
15	CCV2	1	Water	NA	NA	20.1	12.02	09/30/2019	17:42

7B 105393

WORKLIST(Hardcopy Internal Chain)

WorkList Name: CORROSIONITY K5095 WorkList ID: 131454 Department: Wet-Chemistry Date: 09-30-2019 14:53:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/30/2019	Solid	K4659-03	Corrosivity	Cool 4 deg C	PSEG05	N13	JC-02-093019-A	09/30/2019	9045D
09/30/2019	Solid	K4659-06	Corrosivity	Cool 4 deg C	PSEG05	N13	JC-02-093019-B	09/30/2019	9045D
09/30/2019	Solid	K4663-12	Corrosivity	Cool 4 deg C	PSEG05		OK-01-093019	09/30/2019	9045D
09/28/2019	Solid	K4946-16	Corrosivity	Cool 4 deg C	DVIR01	M11	GP-3_(1-3)	09/28/2019	9045D
09/28/2019	Solid	K4946-21	Corrosivity	Cool 4 deg C	DVIR01	M11	GP-5_(10-12)	09/28/2019	9045D
09/27/2019	Solid	K5095-01	Corrosivity	Cool 4 deg C	SEAC01	O41	COMPOSITE-1	09/27/2019	9045D
09/27/2019	Solid	K5095-02	Corrosivity	Cool 4 deg C	SEAC01	O41	COMPOSITE-2	09/27/2019	9045D
09/27/2019	Solid	K5095-03	Corrosivity	Cool 4 deg C	SEAC01	O41	COMPOSITE-3	09/27/2019	9045D

Date/Time 09/30/19 15:00
 Received by: JA
 Relinquished by: [Signature]

Date/Time 09/30/19 17:40
 Received by: [Signature]
 Relinquished by: [Signature]