

TOTAL VOLATILE SOLIDS 160.4

TEMP1 IN: 103 °C 09/30/2019 15:20 **TEMP1 OUT:** 104 °C 10/01/2019 07:10
TEMP2 IN: 104 °C 10/01/2019 08:35 **TEMP2 OUT:** 105 °C 10/01/2019 10:10
TEMP3 IN: 540 °C 10/01/2019 11:40 **TEMP3 OUT:** 540 °C 10/01/2019 12:00
TEMP4 IN: 540 °C 10/01/2019 12:30 **TEMP4 OUT:** 540 °C 10/01/2019 13:10

Run Number: LB105400

SUPERVISOR: apatel

ANALYST: JIGNESH

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB105400BL	84.754	84.754	100	84.754	84.754	84.754	84.754	0	0
2	K4843-36	92.7523	92.7523	100	92.7879	92.7879	92.7549	92.7549	0.033	330

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$