

LB105401


7471B

INSTRUMENT ID : CV1

Linear

 μ Abs.:

59040



A= 0.0000e+000

B= 1.7237e-004

C= -7.4544e-002

Rho= 0.9997517

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.001	-0.001	424	0.000	424				
0.2	0.200	0.227	0.027	1747	0.0 %	1747				
2.5	2.500	2.505	0.005	14965	0.0 %	14965				
5.0	5.000	5.035	0.035	29644	0.0 %	29644				
7.5	7.500	7.332	-0.168	42970	0.0 %	42970				
10.0	10.000	10.102	0.102	59040	0.0 %	59040				

LB105401 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 01 Oct 2019 12:23:56					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	424	-	0.0000	7471B	PPB	01 Oct 2019 12:27:44	S	Std
0.2 - 1	50.2	1747	-	0.2000	7471B	PPB	01 Oct 2019 12:31:46	S	Std
2.5 - 1	52.5	14965	-	2.5000	7471B	PPB	01 Oct 2019 12:34:05	S	Std
5.0 - 1	55.0	29644	-	5.0000	7471B	PPB	01 Oct 2019 12:36:23	S	Std
7.5 - 1	57.5	42970	-	7.5000	7471B	PPB	01 Oct 2019 12:38:41	S	Std
10.0 - 1	510.0	59040	-	10.0000	7471B	PPB	01 Oct 2019 12:40:58	S	Std
ICV35 - 1	ICV35	24245	4.1046	-	7471B	PPB	01 Oct 2019 12:45:25	U	SMPL
ICB35 - 1	ICB35	398	-0.0059	-	7471B	PPB	01 Oct 2019 12:47:41	U	SMPL
CCV34 - 1	CCV34	29493	5.0092	-	7471B	PPB	01 Oct 2019 12:49:58	U	SMPL
CCB34 - 1	CCB34	403	-0.0051	-	7471B	PPB	01 Oct 2019 12:52:14	U	SMPL
CRA - 1	CRA	1756	0.2281	-	7471B	PPB	01 Oct 2019 12:54:30	U	SMPL
HighStd - 1	HighStd	58934	10.0840	-	7471B	PPB	01 Oct 2019 12:56:46	U	SMPL
ChkStd - 1	ChkStd	40292	6.8707	-	7471B	PPB	01 Oct 2019 12:59:03	U	SMPL
PB123527BL - 1	PBS	579	0.0253	-	7471B	PPB	01 Oct 2019 13:01:21	U	SMPL
PB123527BS - 1	LCSS	21539	3.6382	-	7471B	PPB	01 Oct 2019 13:03:38	U	SMPL
K4663-10 - 1	OK-01-093019	1553	0.1932	-	7471B	PPB	01 Oct 2019 13:05:58	U	SMPL
K4946-02 - 1	GP-1_(1-3)	5875	0.9381	-	7471B	PPB	01 Oct 2019 13:08:15	U	SMPL
K4946-03 - 1	GP-1_(10-12)	1514	0.1864	-	7471B	PPB	01 Oct 2019 13:10:33	U	SMPL
K4946-04 - 1	GP-2_(1-3)	10233	1.6893	-	7471B	PPB	01 Oct 2019 13:12:52	U	SMPL
K4946-05 - 1	GP-2_(10-12)	970	0.0927	-	7471B	PPB	01 Oct 2019 13:15:08	U	SMPL
CCV35 - 1	CCV35	29545	5.0182	-	7471B	PPB	01 Oct 2019 13:17:24	U	SMPL
CCB35 - 1	CCB35	397	-0.0061	-	7471B	PPB	01 Oct 2019 13:19:41	U	SMPL
K4946-06 - 1	GP-3_(1-3)	6307	1.0126	-	7471B	PPB	01 Oct 2019 13:21:58	U	SMPL
K4946-07 - 1	GP-3_(10-12)	583	0.0259	-	7471B	PPB	01 Oct 2019 13:24:14	U	SMPL
K4946-08 - 1	GP-4_(1-3)	10875	1.8000	-	7471B	PPB	01 Oct 2019 13:26:31	U	SMPL
K4946-09 - 1	GP-4_(10-12)	552	0.0206	-	7471B	PPB	01 Oct 2019 13:28:47	U	SMPL
K4946-10 - 1	GP-5_(1-3)	3875	0.5934	-	7471B	PPB	01 Oct 2019 13:31:04	U	SMPL
K4946-11 - 1	GP-5_(10-12)	1030	0.1030	-	7471B	PPB	01 Oct 2019 13:33:22	U	SMPL
K5079-01 - 1	TP-1-COMP	2165	0.2986	-	7471B	PPB	01 Oct 2019 13:35:39	U	SMPL
K5079-01DUP - 1	TP-1-COMPDUP	2161	0.2980	-	7471B	PPB	01 Oct 2019 13:37:57	U	SMPL
K5079-01MS - 1	TP-1-COMPMS	22734	3.8442	-	7471B	PPB	01 Oct 2019 13:40:15	U	SMPL
K5079-01MSD - 1	TP-1-COMPMSD	22138	3.7414	-	7471B	PPB	01 Oct 2019 13:42:31	U	SMPL
CCV36 - 1	CCV36	29805	5.0630	-	7471B	PPB	01 Oct 2019 13:44:48	U	SMPL
CCB36 - 1	CCB36	387	-0.0078	-	7471B	PPB	01 Oct 2019 13:47:06	U	SMPL
K5083-01 - 1	20190708-RR-COMPOSITE-8	1582	0.1981	-	7471B	PPB	01 Oct 2019 13:49:22	U	SMPL
K5083-02 - 1	20190707-RR-COMPOSITE-7	5623	0.8947	-	7471B	PPB	01 Oct 2019 13:51:39	U	SMPL
K5083-03 - 1	20190706-RR-COMP-6	8140	1.3286	-	7471B	PPB	01 Oct 2019 13:53:56	U	SMPL
K5083-04 - 1	20190705-RR-COMPOSITE-5	9216	1.5140	-	7471B	PPB	01 Oct 2019 13:56:12	U	SMPL
K5083-05 - 1	20190704-RR-COMPOSITE-4	2761	0.4014	-	7471B	PPB	01 Oct 2019 13:58:28	U	SMPL
K5083-06 - 1	20190703-RR-COMPOSITE-3	3580	0.5425	-	7471B	PPB	01 Oct 2019 14:00:45	U	SMPL
K5083-07 - 1	20190702-RR-COMPOSITE-2	4656	0.7280	-	7471B	PPB	01 Oct 2019 14:03:02	U	SMPL
K5083-08 - 1	20190701-RR-COMPOSITE-1	9085	1.4915	-	7471B	PPB	01 Oct 2019 14:05:20	U	SMPL
K5094-01 - 1	MC-17-COMPOSITE	2031	0.2755	-	7471B	PPB	01 Oct 2019 14:07:38	U	SMPL
K5105-01 - 1	WASTE	3690	0.5615	-	7471B	PPB	01 Oct 2019 14:09:55	U	SMPL
CCV37 - 1	CCV37	29575	5.0234	-	7471B	PPB	01 Oct 2019 14:21:33	U	SMPL
CCB37 - 1	CCB37	395	-0.0065	-	7471B	PPB	01 Oct 2019 14:23:49	U	SMPL
K5079-01LX5 - 1		651	0.0377	-	7471B	PPB	01 Oct 2019 14:26:06	U	SMPL
K5079-01A - 1		22556	3.8135	-	7471B	PPB	01 Oct 2019 14:28:23	U	SMPL
CCV38 - 1	CCV38	29680	5.0415	-	7471B	PPB	01 Oct 2019 14:30:38	U	SMPL
CCB38 - 1	CCB38	379	-0.0092	-	7471B	PPB	01 Oct 2019 14:32:56	U	SMPL