



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105411

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/01/2019	16:20
2	CAL2	1	Water	NA	NA	20.1	7.01	10/01/2019	16:21
3	CAL3	1	Water	NA	NA	20.1	10.03	10/01/2019	16:22
4	ICV	1	Water	NA	NA	20.1	6.99	10/01/2019	16:24
5	CCV1	1	Water	NA	NA	20.1	2.01	10/01/2019	16:25
6	K5134-01	1	Solid	20.02	20	21.5	6.35	10/01/2019	16:26
7	K5134-01DUP	1	Solid	20.03	20	21.6	6.36	10/01/2019	16:27
8	CCV2	1	Water	NA	NA	20.2	12.02	10/01/2019	16:30

VB105411

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5134

WorkList ID : 131514

Department : Wet-Chemistry

Date : 10-01-2019 16:11:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/01/2019	Solid	K5134-01	pH	Cool 4 deg C	TULL02	O51	FK-PS-01-061	10/01/2019	9045D

Date/Time 10.01.19 16:15
Received by: Jn
Relinquished by: CJ

Date/Time 10.01.19 17:00
Received by: CJ
Relinquished by: Jn