



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB105418

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/01/2019	17:15
2	K5095-04		1	9	100.00	>212.0	>212.0	10/01/2019	17:40
3	K5095-04DUP		1	9	100.00	>212.0	>212.0	10/01/2019	18:02
4	K5095-05		1	8	100.00	>212.0	>212.0	10/01/2019	18:32

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 105418

WorkList Name : fp10119

WorkList ID : 131512

Department : Wet-Chemistry

Date : 10-01-2019 15:34:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/11/2019	Water	K5095-04	Flash Point	Cool 4 deg C	SEAC01	O41	COMPOSITE-4	09/27/2019	1010A
10/11/2019	Water	K5095-05	Flash Point	Cool 4 deg C	SEAC01	O41	COMPOSITE-5	09/27/2019	1010A

Date/Time 10/01/19 16:55
 Received by: KS
 Relinquished by: CP

Date/Time 10/01/19 18:40
 Received by: KS
 Relinquished by: KS