



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB105425
Analysis Date: 10/02/2019
BalanceID: WC SC-3
OvenID: EXT OVEN-3

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 10/02/2019
Extraction IN Time: 10:20
Extraction OUT Time: 10:50

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB105425BL	LB105425BL	WATER	1.3	1000	100	3.1569	3.1569	3.02	3.1569	3.1569	0.0000	0
2	LB105425BS	LB105425BS	WATER	1.3	1000	100	3.0147	3.0147	3.01	3.0333	3.0333	0.0186	18.6
3	LB105425BSD	LB105425BSD	WATER	1.3	1000	100	2.6477	2.6477	3.03	2.6665	2.6665	0.0188	18.8
4	K4660-03	TP-1-GROUNDWATER	WATER	1.6	1000	100	3.0064	3.0064	3.04	3.0082	3.0082	0.0018	1.8
5	K5123-04	FE-19B	WATER	1.6	1000	100	2.9768	2.9768	3.03	2.9771	2.9771	0.0003	0.3



QC Batch# LB105425 Test: TPH Analysis Date: 10/02/2019

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2591
pH Paper 0-14	M4240
Sodium Sulfate	EP1987
1:1 HCL	WP74116
Silica Gel	W2546
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP74117
LCSWD	NA	NA
MS/MSD	10.00 ML	WP74118

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP1 :	<u>71 °C</u>	Dessicator Time In1	<u>12:16</u>
1.0000 gram Balance:	<u>1.0004</u>	In Time1:	<u>11:15</u>		
Bal Check Time:	<u>10:35</u>	Out OVEN TEMP1:	<u>71 °C</u>	Dessicator Time Out1	<u>12:40</u>
		Out Time1	<u>12:15</u>		

After Analysis

0.0020 gram Balance:	<u>0.0018</u>	In OVEN TEMP2 :	<u>70 °C</u>	Dessicator Time In2	<u>13:46</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time2:	<u>13:15</u>		
Bal Check Time:	<u>14:40</u>	Out OVEN TEMP2:	<u>70 °C</u>	Dessicator Time Out2	<u>14:15</u>
		Out Time2:	<u>13:45</u>		

105425

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH K5123 WorkList ID : 131557 Department : Wet-Chemistry Date : 10-02-2019 10:06:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/08/2019	Water	K4660-03	K TPH	Conc H2SO4 to	PSEG01	N21	TP-1-GROUNDWATER	10/01/2019	1664A
10/08/2019	Water	K5123-04	B TPH	Conc H2SO4 to	M&MM01	M11	FE-19B	10/01/2019	1664A

Date/Time 10-02-19 10:10
Received by: JS
Relinquished by: cd

Date/Time 10-02-19 11:15
Received by: LD
Relinquished by: 31