


LB105428

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.3941e-004

C= -5.9862e-002

Rho= 0.9999219

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.014	-0.014	192	0.000	192				
0.2	0.200	0.177	-0.023	989	0.0 %	989				
2.5	2.500	2.504	0.004	10708	0.0 %	10708				
5.0	5.000	5.096	0.096	21535	0.0 %	21535				
7.5	7.500	7.448	-0.052	31359	0.0 %	31359				
10.0	10.000	9.989	-0.011	41974	0.0 %	41974				

LB105428 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Oct 2019 11:07:42

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	192	-	0.0000	7470A	PPB	02 Oct 2019 11:40:25	S	Std
0.2 - 1	50.2	989	-	0.2000	7470A	PPB	02 Oct 2019 11:42:34	S	Std
2.5 - 1	82.5	10708	-	2.5000	7470A	PPB	02 Oct 2019 11:44:43	S	Std
5.0 - 1	55.0	21535	-	5.0000	7470A	PPB	02 Oct 2019 11:46:52	S	Std
7.5 - 1	57.5	31359	-	7.5000	7470A	PPB	02 Oct 2019 11:49:00	S	Std
10.0 - 1	510.0	41974	-	10.0000	7470A	PPB	02 Oct 2019 11:51:08	S	Std
ICV37 - 1	ICV37	15731	3.7064	-	7470A	PPB	02 Oct 2019 11:53:52	U	SMPL
ICB37 - 1	ICB37	184	-0.0158	-	7470A	PPB	02 Oct 2019 11:55:59	U	SMPL
CCV42 - 1	CCV42	21478	5.0823	-	7470A	PPB	02 Oct 2019 11:58:05	U	SMPL
CCB42 - 1	CCB42	184	-0.0158	-	7470A	PPB	02 Oct 2019 12:00:12	U	SMPL
CRA - 1	CRA	1018	0.1839	-	7470A	PPB	02 Oct 2019 12:02:19	U	SMPL
HighStd - 1	HighStd	42362	10.0822	-	7470A	PPB	02 Oct 2019 12:04:26	U	SMPL
ChkStd - 1	ChkStd	29436	6.9875	-	7470A	PPB	02 Oct 2019 12:06:33	U	SMPL
PB123574BL - 1	PBW	88	-0.0388	-	7470A	PPB	02 Oct 2019 12:08:39	U	SMPL
PB123574BS - 1	LCSW	16688	3.9355	-	7470A	PPB	02 Oct 2019 12:10:47	U	SMPL
K4659-03 - 1	JC-02-093019-A	229	-0.0050	-	7470A	PPB	02 Oct 2019 12:12:54	U	SMPL
K4659-03DUP - 1	JC-02-093019-ADUP	230	-0.0048	-	7470A	PPB	02 Oct 2019 12:15:02	U	SMPL
K4659-03MS - 1	JC-02-093019-AMS	16954	3.9992	-	7470A	PPB	02 Oct 2019 12:17:10	U	SMPL
K4659-03MSD - 1	JC-02-093019-AMSD	17577	4.1483	-	7470A	PPB	02 Oct 2019 12:19:19	U	SMPL
K4659-06 - 1	JC-02-093019-B	174	-0.0182	-	7470A	PPB	02 Oct 2019 12:21:26	U	SMPL
CCV43 - 1	CCV43	21403	5.0643	-	7470A	PPB	02 Oct 2019 12:23:32	U	SMPL
CCB43 - 1	CCB43	192	-0.0139	-	7470A	PPB	02 Oct 2019 12:25:39	U	SMPL
K4663-12 - 1	OK-01-093019	212	-0.0091	-	7470A	PPB	02 Oct 2019 12:27:46	U	SMPL
K5095-01 - 1	COMPOSITE-1	302	0.0124	-	7470A	PPB	02 Oct 2019 12:29:52	U	SMPL
K5095-02 - 1	COMPOSITE-2	244	-0.0014	-	7470A	PPB	02 Oct 2019 12:31:59	U	SMPL
K5095-03 - 1	COMPOSITE-3	496	0.0589	-	7470A	PPB	02 Oct 2019 12:34:06	U	SMPL
PB123574TB - 1	PB123500TB	276	0.0062	-	7470A	PPB	02 Oct 2019 12:36:13	U	SMPL
K4659-03LX5 - 1		41	-0.0500	-	7470A	PPB	02 Oct 2019 12:38:21	U	SMPL
K4659-03A - 1		17363	4.0971	-	7470A	PPB	02 Oct 2019 12:40:31	U	SMPL
PB123579BL - 1	PBW	155	-0.0228	-	7470A	PPB	02 Oct 2019 12:42:38	U	SMPL
PB123579BS - 1	LCSW	16943	3.9965	-	7470A	PPB	02 Oct 2019 12:44:47	U	SMPL
CCV45 - 1 CCV44	CCV45 CCV44	21509	5.0897	-	7470A	PPB	02 Oct 2019 12:57:27	U	SMPL
CCB45 - 1 CCB44	CCB45 CCB44	70	-0.0431	-	7470A	PPB	02 Oct 2019 12:59:34	U	SMPL
K5111-07 - 1	LOD-MDL-WATER-01-QT4-2019	430	0.0431	-	7470A	PPB	02 Oct 2019 13:27:48	U	SMPL
K5111-08 - 1	LOQ-WATER-02-QT4-2019	1827	0.3775	-	7470A	PPB	02 Oct 2019 13:29:55	U	SMPL
CCV45 - 1	CCV45	21263	5.0308	-	7470A	PPB	02 Oct 2019 13:32:02	U	SMPL
CCB45 - 1	CCB45	171	-0.0189	-	7470A	PPB	02 Oct 2019 13:34:09	U	SMPL

RS 10-02-19