



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB105433

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/02/2019	11:20
2	K4660-04		1	9	100.00	>212.0	>212.0	10/02/2019	11:45
3	K4660-04DUP		1	9	100.00	>212.0	>212.0	10/02/2019	12:20

$$\text{Result} = (\text{Celsius} * 1.8) + 32$$

$$\text{Final Result} = \text{Result} + (760 - \text{Ambient Barometric Pressure}) * 0.06$$

LB 105433

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp10219 WorkList ID : 131540 Department : Wet-Chemistry Date : 10-02-2019 08:49:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/14/2019	Water	K4660-04	Flash Point	Cool 4 deg C	PSEG01	N21	TP-1-GROUNDWATER	10/01/2019	1010A

Date/Time 10/21/19 10:50
 Received by: KS
 Relinquished by: JF

Date/Time 10/21/19 12:25
 Received by: JF
 Relinquished by: KS