

LB105437


7471B

INSTRUMENT ID : CV1

Linear

 μ Abs.:

60326



A= 0.0000e+000

B= 1.6694e-004

C= -5.3974e-002

Rho= 0.9999506

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.024	0.024	468	0.000	468				
0.2	0.200	0.237	0.037	1745	0.0 %	1745				
2.5	2.500	2.425	-0.075	14851	0.0 %	14851				
5.0	5.000	4.988	-0.012	30203	0.0 %	30203				
7.5	7.500	7.509	0.009	45303	0.0 %	45303				
10.0	10.000	10.017	0.017	60326	0.0 %	60326				

LB105437 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 02 Oct 2019 13:54:24					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	468	-	0.0000	7471B	PPB	02 Oct 2019 13:56:03	S	Std
0.2 - 1	50-2	1745	-	0.2000	7471B	PPB	02 Oct 2019 13:58:22	S	Std
2.5 - 1	52-5	14851	-	2.5000	7471B	PPB	02 Oct 2019 14:00:40	S	Std
5.0 - 1	55-0	30203	-	5.0000	7471B	PPB	02 Oct 2019 14:02:58	S	Std
7.5 - 1	57-5	45303	-	7.5000	7471B	PPB	02 Oct 2019 14:05:15	S	Std
10.0 - 1	510-0	60326	-	10.0000	7471B	PPB	02 Oct 2019 14:07:32	S	Std
ICV38 - 1	ICV38	24336	4.0086	-	7471B	PPB	02 Oct 2019 14:12:14	U	SMPL
ICB38 - 1	ICB38	425	0.0170	-	7471B	PPB	02 Oct 2019 14:14:30	U	SMPL
CCV46 - 1	CCV46	30186	4.9852	-	7471B	PPB	02 Oct 2019 14:16:46	U	SMPL
CCB46 - 1	CCB46	416	0.0155	-	7471B	PPB	02 Oct 2019 14:19:02	U	SMPL
CRA - 1	CRA	1728	0.2345	-	7471B	PPB	02 Oct 2019 14:21:18	U	SMPL
HighStd - 1	HighStd	59772	9.9241	-	7471B	PPB	02 Oct 2019 14:23:34	U	SMPL
ChkStd - 1	ChkStd	42290	7.0057	-	7471B	PPB	02 Oct 2019 14:25:50	U	SMPL
PB123585BL - 1	PBS	561	0.0397	-	7471B	PPB	02 Oct 2019 14:28:06	U	SMPL
PB123585BS - 1	LCSS	22103	3.6358	-	7471B	PPB	02 Oct 2019 14:30:24	U	SMPL
K4659-01 - 1	JC-02-093019-A	9626	1.5530	-	7471B	PPB	02 Oct 2019 14:32:40	U	SMPL
K4659-01DUP - 1	JC-02-093019-ADUP	9722	1.5690	-	7471B	PPB	02 Oct 2019 14:34:58	U	SMPL
K4659-01MS - 1	JC-02-093019-AMS	29133	4.8094	-	7471B	PPB	02 Oct 2019 14:37:15	U	SMPL
K4659-01MSD - 1	JC-02-093019-AMSD	31756	5.2472	-	7471B	PPB	02 Oct 2019 14:39:34	U	SMPL
K4659-04 - 1	JC-02-093019-B	8890	1.4301	-	7471B	PPB	02 Oct 2019 14:41:49	U	SMPL
CCV47 - 1	CCV47	30159	4.9806	-	7471B	PPB	02 Oct 2019 14:44:07	U	SMPL
CCB47 - 1	CCB47	507	0.0307	-	7471B	PPB	02 Oct 2019 14:46:23	U	SMPL
K4660-01 - 1	TP-1	2617	0.3829	-	7471B	PPB	02 Oct 2019 14:48:39	U	SMPL
K5127-01 - 1	20190772-SS-COMP-12	11469	1.8606	-	7471B	PPB	02 Oct 2019 14:50:56	U	SMPL
K5127-02 - 1	20190762-SS-COMP-2	13807	2.2509	-	7471B	PPB	02 Oct 2019 14:53:12	U	SMPL
K5128-01 - 1	SB-1A	41936	6.9467	-	7471B	PPB	02 Oct 2019 14:55:28	U	SMPL
K5133-01 - 1	S-1	1463	0.1903	-	7471B	PPB	02 Oct 2019 14:57:45	U	SMPL
K5134-01 - 1	FK-PS-01-061	5453	0.8563	-	7471B	PPB	02 Oct 2019 15:00:03	U	SMPL
K5137-01 - 1	FK-GF-02-047-A	8363	1.3421	-	7471B	PPB	02 Oct 2019 15:02:20	U	SMPL
K5137-03 - 1	FK-GF-02-047-B	12235	1.9885	-	7471B	PPB	02 Oct 2019 15:04:39	U	SMPL
K5144-01 - 1	COMP-1	31070	5.1327	-	7471B	PPB	02 Oct 2019 15:06:56	U	SMPL
K5148-01 - 1	OR-02-100119	7188	1.1460	-	7471B	PPB	02 Oct 2019 15:09:13	U	SMPL
CCV48 - 1	CCV48	30223	4.9913	-	7471B	PPB	02 Oct 2019 15:11:30	U	SMPL
CCB48 - 1	CCB48	527	0.0340	-	7471B	PPB	02 Oct 2019 15:13:46	U	SMPL
K4659-01LX5 - 1		2212	0.3153	-	7471B	PPB	02 Oct 2019 15:16:02	U	SMPL
K4659-01A - 1		30421	5.0244	-	7471B	PPB	02 Oct 2019 15:18:18	U	SMPL
CCV49 - 1	CCV49	30326	5.0085	-	7471B	PPB	02 Oct 2019 15:20:34	U	SMPL
CCB49 - 1	CCB49	532	0.0348	-	7471B	PPB	02 Oct 2019 15:22:51	U	SMPL