



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 10/3/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:26
In Date: 10/02/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:50
Out Date: 10/03/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB105438

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
K5140-01	HR-05-100219	21	1.15	9.91	8.97	89.3	
K5140-02	HR-05-100219-1	22	1.15	9.75	9.07	92.1	
K5143-01	NB-08-100219	23	1.13	9.74	9.42	96.3	
K5143-02	NB-08-100219	24	1.14	9.74	9.22	94.0	
K5143-03	NB-08-100219	25	1.13	9.72	9.1	92.8	
K5155-01	STOCKPILE	16	1.13	9.91	8.75	86.8	
K5156-01	PSEG-GAS	1	1.15	9.61	6.76	66.3	
K5156-04	EAST-ACCESS-SP	2	1.16	9.65	8.98	92.1	
K5158-01	ROLL-OFF	17	1.14	9.68	9.39	96.6	
K5158-02	ROLL-OFF-1	18	1.14	9.68	9.39	96.6	
K5158-03	ROLL-OFF-2	19	1.14	9.68	9.39	96.6	
K5160-01	CSA-3-1	3	1.16	9.83	9.1	91.6	
K5160-03	CSA-3-2	4	1.15	9.66	9.08	93.2	
K5165-01	NBP-2019	5	1.14	9.55	9.36	97.7	
K5168-01	20190763-SS-COMP-3	6	1.14	9.65	6.16	59.0	
K5168-02	20190764-SS-COMP-4	7	1.15	9.7	6.17	58.7	
K5168-03	20190769-SS-COMP-9	8	1.13	9.79	5.76	53.5	
K5168-04	2019073-SS-COMP-13	9	1.14	9.9	6.3	58.9	
K5168-05	2019071-SS-COMP-11	10	1.13	9.76	6.27	59.6	
K5168-06	2019070-SS-COMP-10	11	1.15	9.84	6.16	57.7	
K5171-01	ETGI-341	20	1.13	9.97	9.6	95.8	
K5181-01	SOUTH-AVENUE-1-ABCD	12	1.13	9.82	8.06	79.7	
K5181-02	SOUTH-AVENUE-2-ABCD	13	1.15	9.84	8.63	86.1	
K5181-03	SOUTH-AVENUE-3-ABCD	14	1.13	9.55	8.2	84.0	
K5181-04	SOUTH-AVENUE-4-ABCD	15	1.14	9.91	8.32	81.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$