

OVEN TEMP1: 98 °C

TEMP: 104 °C

DATE & TIME IN: 10/02/2019 16:20 OUT: 10/03/2019 07:30

OVEN TEMP2: 104 °C

TEMP: 104 °C

DATE & TIME IN: 10/03/2019 08:10 OUT: 10/03/2019 10:35

Run Number: LB105441

BalanceID: WC SC-3

OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB105441BL	LB105441BL	71.5398	71.5398	100	71.5398	71.5398	71.5398	0	0
2	K5131-11	20190825-COMPOSITE-A	76.9191	76.9191	100	77.0291	77.0291	77.0291	0.11	1100
3	K5131-12	20190826-COMPOSITE-B	76.7409	76.7409	100	76.8523	76.8523	76.8523	0.1114	1114
4	K5131-13	20190827-COMPOSITE-C	79.5102	79.5102	100	79.6259	79.6259	79.6259	0.1157	1157
5	K5131-14	20190828-COMPOSITE-D	78.4863	78.4863	100	78.6161	78.6161	78.6161	0.1298	1298
6	K5131-15	20190829-COMPOSITE-E	83.3628	83.3628	100	83.4708	83.4708	83.4708	0.108	1080
7	K5131-16	20190830-COMPOSITE-F	82.1373	82.1373	100	82.2655	82.2655	82.2655	0.1282	1282
8	K5131-17	20190831-COMPOSITE-G	82.4736	82.4736	100	82.6039	82.6039	82.6039	0.1303	1303
9	K5131-18	20190832-COMPOSITE-H	79.1455	79.1455	100	79.2638	79.2638	79.2638	0.1183	1183
10	K5131-19	20190833-COMPOSITE-I	84.7555	84.7555	100	84.8691	84.8691	84.8691	0.1136	1136
11	K5131-20	20190834-COMPOSITE-J	92.7542	92.7542	100	92.8643	92.8643	92.8643	0.1101	1101
12	K5131-20DUP	20190834-COMPOSITE-JDUP	90.0144	90.0144	100	90.1249	90.1249	90.1249	0.1105	1105

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

WORKLIST(Hardcopy Internal Chain)

2B 10544)

WorkList Name : ASTM TS K5123

WorkList ID : 131581

Department : Wet-Chemistry

Date : 10-02-2019 15:33:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/04/2019	Solid	K5131-11	ASTM TS	Cool 4 deg C	AQUA04	O62	20190825-COMPOSITE-A	09/27/2019	SM2540 B
10/04/2019	Solid	K5131-12	ASTM TS	Cool 4 deg C	AQUA04	O62	20190826-COMPOSITE-B	09/27/2019	SM2540 B
10/04/2019	Solid	K5131-13	ASTM TS	Cool 4 deg C	AQUA04	O62	20190827-COMPOSITE-C	09/27/2019	SM2540 B
10/04/2019	Solid	K5131-14	ASTM TS	Cool 4 deg C	AQUA04	O62	20190828-COMPOSITE-D	09/27/2019	SM2540 B
10/04/2019	Solid	K5131-15	ASTM TS	Cool 4 deg C	AQUA04	O62	20190829-COMPOSITE-E	09/27/2019	SM2540 B
10/04/2019	Solid	K5131-16	ASTM TS	Cool 4 deg C	AQUA04	O62	20190830-COMPOSITE-F	09/27/2019	SM2540 B
10/04/2019	Solid	K5131-17	ASTM TS	Cool 4 deg C	AQUA04	O62	20190831-COMPOSITE-G	09/27/2019	SM2540 B
10/04/2019	Solid	K5131-18	ASTM TS	Cool 4 deg C	AQUA04	O62	20190832-COMPOSITE-H	09/27/2019	SM2540 B
10/04/2019	Solid	K5131-19	ASTM TS	Cool 4 deg C	AQUA04	O62	20190833-COMPOSITE-I	09/27/2019	SM2540 B
10/04/2019	Solid	K5131-20	ASTM TS	Cool 4 deg C	AQUA04	O62	20190834-COMPOSITE-J	09/27/2019	SM2540 B

Date/Time 10-02-19 15:40
 Received by: Jg
 Relinquished by: Jg

Date/Time 10-02-19 17:00
 Received by: Jg
 Relinquished by: Jg