



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB105448

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/02/2019	16:00
2	CAL2	1	Water	NA	NA	20.1	7.01	10/02/2019	16:01
3	CAL3	1	Water	NA	NA	20.1	10.03	10/02/2019	16:02
4	ICV	1	Water	NA	NA	20.1	6.99	10/02/2019	16:05
5	CCV1	1	Water	NA	NA	20.1	2.02	10/02/2019	16:06
6	K5140-04	1	Solid	20.02	20	22.9	9.26	10/02/2019	16:07
7	K5140-04DUP	1	Solid	20.03	20	22.8	9.27	10/02/2019	16:08
8	K5143-04	1	Solid	20.03	20	22.9	6.52	10/02/2019	16:10
9	K5156-01	1	Solid	20.02	20	23.2	7.04	10/02/2019	16:11
10	K5156-04	1	Solid	20.03	20	23.4	9.86	10/02/2019	16:12
11	CCV2	1	Water	NA	NA	20.1	12.01	10/02/2019	16:15

LB 105448

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY K5156 WorkList ID : 131580 Department : Wet-Chemistry Date : 10-02-2019 14:30:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/02/2019	Solid	K5140-04	Corrosivity	Cool 4 deg C	PSEG05	N21	HR-05-100219	10/02/2019	9045D
10/02/2019	Solid	K5143-04	Corrosivity	Cool 4 deg C	PSEG05	N21	NB-08-100219	10/02/2019	9045D
10/02/2019	Solid	K5156-01	Corrosivity	Cool 4 deg C	UNIO02	O61	PSEG-GAS	10/02/2019	9045D
10/02/2019	Solid	K5156-04	Corrosivity	Cool 4 deg C	UNIO02	O61	EAST-ACCESS-SP	10/02/2019	9045D

Date/Time 10-02-19 14:35
 Received by: SA
 Relinquished by: SA

Date/Time 10-02-19 18:10
 Received by: SA
 Relinquished by: SA