



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB105449

Slope : 98.6

pH Meter ID : WC PH METER

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/02/2019	16:00
2	CAL2	1	Water	NA	NA	20.1	7.01	10/02/2019	16:01
3	CAL3	1	Water	NA	NA	20.1	10.03	10/02/2019	16:02
4	ICV	1	Water	NA	NA	20.1	6.99	10/02/2019	16:05
5	CCV1	1	Water	NA	NA	20.1	2.02	10/02/2019	16:45
6	K5185-01	1	Water	NA	NA	21.0	7.48	10/02/2019	16:46
7	K5185-01DUP	1	Water	NA	NA	21.1	7.49	10/02/2019	16:47
8	CCV2	1	Water	NA	NA	20.1	12.02	10/02/2019	16:48

10/05/19

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY K5185 W WorkList ID : 131582 Department : Wet-Chemistry Date : 10-02-2019 16:40:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/02/2019	Water	K5185-01	Corrosivity	Cool 4 deg C	AREC01	O32	WC-1	10/02/2019	9040C

Date/Time 10-02-19 16:45
 Received by: SA
 Relinquished by: SA

Date/Time 10-02-19 18:20
 Received by: SA
 Relinquished by: SA