



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105450

Slope : 98.6

pH Meter ID : WC PH METER

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/02/2019	16:00
2	CAL2	1	Water	NA	NA	20.1	7.01	10/02/2019	16:01
3	CAL3	1	Water	NA	NA	20.1	10.03	10/02/2019	16:02
4	ICV	1	Water	NA	NA	20.1	6.99	10/02/2019	16:05
5	CCV1	1	Water	NA	NA	20.1	2.02	10/02/2019	17:20
6	K5176-11	1	Water	NA	NA	23.9	7.40	10/02/2019	17:21
7	K5176-11DUP	1	Water	NA	NA	23.8	7.41	10/02/2019	17:22
8	K5176-12	1	Water	NA	NA	24.4	8.22	10/02/2019	17:23
9	K5176-13	1	Water	NA	NA	24.5	7.55	10/02/2019	17:25
10	K5176-14	1	Water	NA	NA	24.4	7.76	10/02/2019	17:30
11	K5176-20	1	Water	NA	NA	24.7	7.54	10/02/2019	17:32
12	K5176-22	1	Water	NA	NA	24.7	7.77	10/02/2019	17:35
13	K5176-23	1	Water	NA	NA	24.7	7.69	10/02/2019	17:36
14	K5180-01	1	Water	NA	NA	24.0	5.23	10/02/2019	17:38
15	K5180-02	1	Water	NA	NA	24.1	5.23	10/02/2019	17:40
16	CCV2	1	Water	NA	NA	20.1	12.01	10/02/2019	17:42

VB106450

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5176 WorkList ID : 131583 Department : Wet-Chemistry Date : 10-02-2019 17:18:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/02/2019	Water	K5176-11 ✓	pH	Cool 4 deg C	PSEG01	N21	30464	10/02/2019	9040C
10/02/2019	Water	K5176-12 ✓	pH	Cool 4 deg C	PSEG01	N21	30465	10/02/2019	9040C
10/02/2019	Water	K5176-13 ✓	pH	Cool 4 deg C	PSEG01	N21	30466	10/02/2019	9040C
10/02/2019	Water	K5176-14 ✓	pH	Cool 4 deg C	PSEG01	N21	30467	10/02/2019	9040C
10/02/2019	Water	K5176-20 ✓	pH	Cool 4 deg C	PSEG01	N21	30459	10/02/2019	9040C
10/02/2019	Water	K5176-22 ✓	pH	Cool 4 deg C	PSEG01	N21	30460	10/02/2019	9040C
10/02/2019	Water	K5176-23 ✓	pH	Cool 4 deg C	PSEG01	N21	30461	10/02/2019	9040C
10/02/2019	Water	K5180-01 ✓	pH	Cool 4 deg C	PSEG01	N21	30474	10/02/2019	9040C
10/02/2019	Water	K5180-02 ✓	pH	Cool 4 deg C	PSEG01	N21	40092	10/02/2019	9040C

Date/Time 10.02.19 17:20
 Received by: fn
 Relinquished by: cd

Date/Time 10.02.19 18:30
 Received by: cd
 Relinquished by: fn