



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 10/4/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 14:53
In Date: 10/03/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:36
Out Date: 10/04/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB105459

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K5179-01	MBEP83	1	1.15	9.8	7.11	68.9	
K5179-02	MBEP84	2	1.12	9.95	7.73	74.9	
K5179-03	MBEP84D	3	1.12	9.95	7.73	74.9	
K5179-04	MBEP84S	4	1.12	9.95	7.73	74.9	
K5179-05	MBEP85	5	1.13	9.81	7.22	70.2	
K5179-06	MBEP86	6	1.12	9.88	8.23	81.2	
K5179-07	MBEP87	7	1.12	9.64	7.65	76.6	
K5179-08	MBEP88	8	1.13	9.77	8.06	80.2	
K5179-09	MBEP89	9	1.15	9.92	6.03	55.6	
K5179-10	MBEPF7	10	1.13	9.98	7.92	76.7	
K5179-11	MBEPF8	11	1.15	9.78	6.26	59.2	
K5179-12	MBEPF9	12	1.13	9.73	6.03	57.0	
K5179-13	MBEPG0	13	1.13	9.81	7.33	71.4	
K5179-14	MBEPG1	14	1.15	9.57	7.7	77.8	
K5179-15	MBEPG2	15	1.14	9.85	7.14	68.9	
K5179-16	MBEPG3	16	1.14	9.8	7.9	78.1	
K5179-17	MBEPG4	17	1.15	9.85	7.58	73.9	
K5179-18	MBEP90	18	1.16	9.91	6.25	58.2	
K5179-19	MBEP91	19	1.13	9.7	7.00	68.5	
K5179-20	MBEP92	20	1.14	9.96	7.88	76.4	
K5179-21	MBEPG5	21	1.14	9.94	7.75	75.1	
K5179-22	MBEPG6	22	1.13	9.91	7.5	72.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$