


LB105461

ISM02.4

INSTRUMENT ID: CU1



Linear

A= 0.0000e+000

B= 2.4940e-004

C= -3.9579e-003

Rho= 0.9998170

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	0.063	0.063	267	0.000	267				
0.05	0.050									
0.20	0.200	0.201	0.001	822	0.0 %	822				
2.50	2.500	2.450	-0.050	9841	0.0 %	9841				
5.00	5.000	4.897	-0.103	19652	0.0 %	19652				
7.50	7.500	7.612	0.112	30535	0.0 %	30535				
10.0	10.000	9.977	-0.023	40020	0.0 %	40020				

%D
—
1
2
2
1
0

RS 10.03.19

LB105461
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	267	-		0 ISM02.4	PPB	10/3/2019 10:58	Std
	0.2 S0.2	822	-		0.2 ISM02.4	PPB	10/3/2019 11:00	Std
	2.5 S2.5	9841	-		2.5 ISM02.4	PPB	10/3/2019 11:02	Std
	5 S5.0	19652	-		5 ISM02.4	PPB	10/3/2019 11:04	Std
	7.5 S7.5	30535	-		7.5 ISM02.4	PPB	10/3/2019 11:07	Std
	10 S10.0	40020	-		10 ISM02.4	PPB	10/3/2019 11:09	Std
ICV	ICV	15901	3.9618 -		ISM02.4	PPB	10/3/2019 11:12	SMPL
ICB	ICB	169	0.0382 -		ISM02.4	PPB	10/3/2019 11:14	SMPL
CCV050	CCV050	19457	4.8487 -		ISM02.4	PPB	10/3/2019 11:16	SMPL
CCB050	CCB050	167	0.0377 -		ISM02.4	PPB	10/3/2019 11:19	SMPL
PB123609BL	PBW609	190	0.0434 -		ISM02.4	PPB	10/3/2019 11:21	SMPL
K5068-01	MDBAF7	410	0.0983 -		ISM02.4	PPB	10/3/2019 11:23	SMPL
K5068-02	MDBAF8	325	0.0771 -		ISM02.4	PPB	10/3/2019 11:26	SMPL
K5068-03	MDBAF9	273	0.0641 -		ISM02.4	PPB	10/3/2019 11:28	SMPL
K5068-04	MDBAF9D	268	0.0629 -		ISM02.4	PPB	10/3/2019 11:30	SMPL
K5068-05	MDBAF9S	5129	1.2752 -		ISM02.4	PPB	10/3/2019 11:32	SMPL
K5068-06	MDBAG0	202	0.0464 -		ISM02.4	PPB	10/3/2019 11:35	SMPL
K5068-07	MDBAG2	210	0.0484 -		ISM02.4	PPB	10/3/2019 11:37	SMPL
K5068-08	MDBAG4	1682	0.4155 -		ISM02.4	PPB	10/3/2019 11:39	SMPL
PB123610BL	PBW610	228	0.0529 -		ISM02.4	PPB	10/3/2019 11:42	SMPL
K5069-01	MDBAF3	1975	0.4886 -		ISM02.4	PPB	10/3/2019 11:44	SMPL
K5069-02	MDBAF4	253	0.0591 -		ISM02.4	PPB	10/3/2019 11:46	SMPL
K5069-03	MDBAF5	221	0.0512 -		ISM02.4	PPB	10/3/2019 11:48	SMPL
K5069-04	MDBAK8	973	0.2387 -		ISM02.4	PPB	10/3/2019 11:51	SMPL
K5069-05	MDBAK8D	964	0.2365 -		ISM02.4	PPB	10/3/2019 11:53	SMPL
K5069-06	MDBAK8S	5880	1.4625 -		ISM02.4	PPB	10/3/2019 11:55	SMPL
PB123609TB	LEB609	277	0.0651 -		ISM02.4	PPB	10/3/2019 11:57	SMPL
PB123610TB	LEB610	261	0.0611 -		ISM02.4	PPB	10/3/2019 12:00	SMPL
CCV051	CCV051	19462	4.8499 -		ISM02.4	PPB	10/3/2019 12:02	SMPL
CCB051	CCB051	166	0.0374 -		ISM02.4	PPB	10/3/2019 12:04	SMPL