

7470A


LB105464

INSTRUMENT ID : CV1

Linear

 μ Abs.:

40347



A= 0.0000e+000

B= 2.4858e-004

C= 6.6288e-003

Rho= 0.9999526

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.022	0.022	63	0.000	63				
0.2	0.200	0.222	0.022	867	0.0 %	867				
2.5	2.500	2.490	-0.010	9992	0.0 %	9992				
5.0	5.000	4.927	-0.073	19793	0.0 %	19793				
7.5	7.500	7.502	0.002	30155	0.0 %	30155				
10.0	10.000	10.036	0.036	40347	0.0 %	40347				

LB105464 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 03 Oct 2019 12:30:20					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	63	-	0.0000	7470A	PPB	03 Oct 2019 12:44:47	S	Std
0.2 - 1	50.2	867	-	0.2000	7470A	PPB	03 Oct 2019 12:46:57	S	Std
2.5 - 1	52.5	9992	-	2.5000	7470A	PPB	03 Oct 2019 12:49:06	S	Std
5.0 - 1	55.0	19793	-	5.0000	7470A	PPB	03 Oct 2019 12:51:14	S	Std
7.5 - 1	57.5	30155	-	7.5000	7470A	PPB	03 Oct 2019 12:53:23	S	Std
10.0 - 1	510.0	40347	-	10.0000	7470A	PPB	03 Oct 2019 12:55:31	S	Std
ICV40 - 1	ICV40	16227	4.0403	-	7470A	PPB	03 Oct 2019 13:01:52	U	SMPL
ICB40 - 1	ICB40	13	0.0099	-	7470A	PPB	03 Oct 2019 13:03:58	U	SMPL
CCV52 - 1	CCV52	19877	4.9476	-	7470A	PPB	03 Oct 2019 13:06:05	U	SMPL
CCB52 - 1	CCB52	-5	0.0054	-	7470A	PPB	03 Oct 2019 13:08:12	U	SMPL
CRA - 1	CRA	846	0.2169	-	7470A	PPB	03 Oct 2019 13:10:19	U	SMPL
HighStd - 1	HighStd	40326	10.0307	-	7470A	PPB	03 Oct 2019 13:12:25	U	SMPL
ChkStd - 1	ChkStd	27800	6.9171	-	7470A	PPB	03 Oct 2019 13:14:32	U	SMPL
PB123616BL - 1	PBW	311	0.0839	-	7470A	PPB	03 Oct 2019 13:16:40	U	SMPL
PB123616BS - 1	LCSW	17347	4.3187	-	7470A	PPB	03 Oct 2019 13:18:48	U	SMPL
K5140-04 - 1	HR-05-100219	14	0.0101	-	7470A	PPB	03 Oct 2019 13:20:55	U	SMPL
K5140-04DUP - 1	HR-05-100219DUP	25	0.0128	-	7470A	PPB	03 Oct 2019 13:23:04	U	SMPL
K5140-04MS - 1	HR-05-100219MS	16691	4.1556	-	7470A	PPB	03 Oct 2019 13:25:12	U	SMPL
K5140-04MSD - 1	HR-05-100219MSD	16069	4.0010	-	7470A	PPB	03 Oct 2019 13:27:22	U	SMPL
K5143-04 - 1	NB-08-100219	18	0.0111	-	7470A	PPB	03 Oct 2019 13:29:28	U	SMPL
CCV53 - 1	CCV53	19881	4.9486	-	7470A	PPB	03 Oct 2019 13:31:35	U	SMPL
CCB53 - 1	CCB53	2	0.0071	-	7470A	PPB	03 Oct 2019 13:33:42	U	SMPL
K5156-02 - 1	PSEG-GAS	61	0.0218	-	7470A	PPB	03 Oct 2019 13:35:49	U	SMPL
K5156-05 - 1	EAST-ACCESS-SP	65	0.0228	-	7470A	PPB	03 Oct 2019 13:37:56	U	SMPL
K5160-02 - 1	CSA-3-1	53	0.0198	-	7470A	PPB	03 Oct 2019 13:40:03	U	SMPL
K5160-04 - 1	CSA-3-2	64	0.0225	-	7470A	PPB	03 Oct 2019 13:42:11	U	SMPL
PB123617BL - 1	PBW	58	0.0210	-	7470A	PPB	03 Oct 2019 13:44:19	U	SMPL
PB123617BS - 1	LCSW	15923	3.9647	-	7470A	PPB	03 Oct 2019 13:46:26	U	SMPL
K4660-02 - 1	TP-1	388	0.1031	-	7470A	PPB	03 Oct 2019 13:48:35	U	SMPL
K4660-02DUP - 1	TP-1DUP	369	0.0984	-	7470A	PPB	03 Oct 2019 13:50:43	U	SMPL
K4660-02MS - 1	TP-1MS	16743	4.1686	-	7470A	PPB	03 Oct 2019 13:52:52	U	SMPL
K4660-02MSD - 1	TP-1MSD	13454	3.3510	-	7470A	PPB	03 Oct 2019 13:54:59	U	SMPL
CCV54 - 1	CCV54	19823	4.9342	-	7470A	PPB	03 Oct 2019 13:57:06	U	SMPL
CCB54 - 1	CCB54	-17	0.0024	-	7470A	PPB	03 Oct 2019 13:59:14	U	SMPL
K5131-11 - 1	20190825-COMPOSITE-A	341	0.0914	-	7470A	PPB	03 Oct 2019 14:01:21	U	SMPL
K5131-12 - 1	20190826-COMPOSITE-B	370	0.0986	-	7470A	PPB	03 Oct 2019 14:03:27	U	SMPL
K5131-13 - 1	20190827-COMPOSITE-C	358	0.0956	-	7470A	PPB	03 Oct 2019 14:05:35	U	SMPL
K5131-14 - 1	20190828-COMPOSITE-D	353	0.0944	-	7470A	PPB	03 Oct 2019 14:07:42	U	SMPL
K5131-15 - 1	20190829-COMPOSITE-E	349	0.0934	-	7470A	PPB	03 Oct 2019 14:09:50	U	SMPL
K5131-16 - 1	20190830-COMPOSITE-F	346	0.0926	-	7470A	PPB	03 Oct 2019 14:11:57	U	SMPL
K5131-17 - 1	20190831-COMPOSITE-G	331	0.0889	-	7470A	PPB	03 Oct 2019 14:14:06	U	SMPL
K5131-18 - 1	20190832-COMPOSITE-H	318	0.0857	-	7470A	PPB	03 Oct 2019 14:16:14	U	SMPL
K5131-19 - 1	20190833-COMPOSITE-I	263	0.0720	-	7470A	PPB	03 Oct 2019 14:18:23	U	SMPL
K5131-20 - 1	20190834-COMPOSITE-J	256	0.0703	-	7470A	PPB	03 Oct 2019 14:20:30	U	SMPL
CCV55 - 1	CCV55	19854	4.9419	-	7470A	PPB	03 Oct 2019 14:22:37	U	SMPL
CCB55 - 1	CCB55	-11	0.0039	-	7470A	PPB	03 Oct 2019 14:24:45	U	SMPL
K5148-04 - 1	OR-02-100119	221	0.0616	-	7470A	PPB	03 Oct 2019 14:26:52	U	SMPL
PB123618BL - 1	PBW	546	0.1424	-	7470A	PPB	03 Oct 2019 14:29:00	U	SMPL
PB123618BS - 1	LCSW	17460	4.3468	-	7470A	PPB	03 Oct 2019 14:31:06	U	SMPL
K4660-03 - 1	TP-1-GROUNDWATER	50808	12.6363	-	7470A	PPB	03 Oct 2019 14:33:14	U	SMPL
K5154-02 - 1	FB-10012019	226	0.0628	-	7470A	PPB	03 Oct 2019 14:44:14	U	SMPL
K5154-03 - 1	MW-106S	530	0.1384	-	7470A	PPB	03 Oct 2019 14:46:22	U	SMPL
K5154-04 - 1	MW-106I	506	0.1324	-	7470A	PPB	03 Oct 2019 14:48:29	U	SMPL
K5154-05 - 1	MW-106D	457	0.1202	-	7470A	PPB	03 Oct 2019 14:50:38	U	SMPL
K5154-06 - 1	MW-107	450	0.1185	-	7470A	PPB	03 Oct 2019 14:52:47	U	SMPL
K5154-10 - 1	MW-103-S	106	0.0330	-	7470A	PPB	03 Oct 2019 14:54:55	U	SMPL
CCV56 - 1	CCV56	19716	4.9076	-	7470A	PPB	03 Oct 2019 14:57:02	U	SMPL
CCB56 - 1	CCB56	-10	0.0041	-	7470A	PPB	03 Oct 2019 14:59:10	U	SMPL
K5154-11 - 1	MW-103-I	27	0.0133	-	7470A	PPB	03 Oct 2019 15:01:18	U	SMPL
K5154-11DUP - 1	MW-103-IDUP	29	0.0138	-	7470A	PPB	03 Oct 2019 15:03:25	U	SMPL
K5154-12 - 1	K5154-11MS	17615	4.3853	-	7470A	PPB	03 Oct 2019 15:05:32	U	SMPL
K5154-13 - 1	K5154-11MSD	17160	4.2722	-	7470A	PPB	03 Oct 2019 15:07:40	U	SMPL
K5154-14 - 1	MW-103-D	-12	0.0036	-	7470A	PPB	03 Oct 2019 15:09:49	U	SMPL
K5154-15 - 1	MW-105-S	85	0.0278	-	7470A	PPB	03 Oct 2019 15:11:57	U	SMPL
K5154-16 - 1	MW-105-I	55	0.0203	-	7470A	PPB	03 Oct 2019 15:14:05	U	SMPL
K5154-17 - 1	MW-105-D	21	0.0118	-	7470A	PPB	03 Oct 2019 15:16:13	U	SMPL
K5154-18 - 1	MW-104-I	30	0.0141	-	7470A	PPB	03 Oct 2019 15:18:22	U	SMPL
K5154-19 - 1	MW-104-D	109	0.0337	-	7470A	PPB	03 Oct 2019 15:20:29	U	SMPL
CCV57 - 1	CCV57	20085	4.9993	-	7470A	PPB	03 Oct 2019 15:22:36	U	SMPL
CCB57 - 1	CCB57	0	0.0066	-	7470A	PPB	03 Oct 2019 15:24:44	U	SMPL
K5176-10 - 1	30464-67	1789	0.4513	-	7470A	PPB	03 Oct 2019 15:26:51	U	SMPL
K5176-19 - 1	30459-61	3295	0.8257	-	7470A	PPB	03 Oct 2019 15:28:59	U	SMPL
K5180-02 - 1	40092	671	0.1734	-	7470A	PPB	03 Oct 2019 15:31:06	U	SMPL
K5185-01 - 1	WC-1	-176	-0.0371	-	7470A	PPB	03 Oct 2019 15:33:14	U	SMPL
PB123616TB - 1	PB123586TB	15	0.0104	-	7470A	PPB	03 Oct 2019 15:35:22	U	SMPL
PB123617TB - 1	PB123549TB	338	0.0906	-	7470A	PPB	03 Oct 2019 15:37:30	U	SMPL
K5140-04LX5 - 1		14	0.0101	-	7470A	PPB	03 Oct 2019 15:39:39	U	SMPL
K5140-04A - 1		16414	4.0868	-	7470A	PPB	03 Oct 2019 15:41:47	U	SMPL
K4660-02LX5 - 1		89	0.0288	-	7470A	PPB	03 Oct 2019 15:43:56	U	SMPL
K4660-02A - 1		13535	3.3711	-	7470A	PPB	03 Oct 2019 15:46:04	U	SMPL
CCV58 - 1	CCV58	19804	4.9294	-	7470A	PPB	03 Oct 2019 15:48:12	U	SMPL

LB105464 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Oct 2019 12:30:20

Sample ID	Extended ID	μ Abs.	Conc.	Strnd Conc	Method	Units	Date	Type	Type
CCB58 - 1	CCB58	4	0.0076		-7470A	PPB	03 Oct 2019 15:50:20	U	SMPL
K5154-11LX5 - 1		5	0.0079		-7470A	PPB	03 Oct 2019 15:52:28	U	SMPL
K5154-11A - 1		17013	4.2357		-7470A	PPB	03 Oct 2019 15:54:35	U	SMPL
K4660-03DLX2 - 1		24790	6.1689		-7470A	PPB	03 Oct 2019 15:56:44	U	SMPL
CCV59 - 1	CCV59	19751	4.9163		-7470A	PPB	03 Oct 2019 15:58:52	U	SMPL
CCB59 - 1	CCB59	159	0.0462		-7470A	PPB	03 Oct 2019 16:00:59	U	SMPL