



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB105467

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/03/2019	13:20
2	K5176-10		1	9	100.00	>212.0	>212.0	10/03/2019	13:45
3	K5176-10DUP		1	9	100.00	>212.0	>212.0	10/03/2019	14:10
4	K5176-19		1	7	100.00	>212.0	>212.0	10/03/2019	14:35
5	K5180-01		1	8	100.00	>212.0	>212.0	10/03/2019	14:50
6	K5180-02		1	9	100.00	>212.0	>212.0	10/03/2019	15:22
7	K5185-01		1	8	100.00	>212.0	>212.0	10/03/2019	15:48

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 105467

WorkList Name : fp10319 WorkList ID : 131621 Department : Wet-Chemistry Date : 10-03-2019 11:24:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/09/2019	Water	K5176-10 <u>P</u>	Flash Point	Cool 4 deg C	PSEG01	N21	30464-67	10/02/2019	1010A
10/09/2019	Water	K5176-19 <u>E</u>	Flash Point	Cool 4 deg C	PSEG01	N21	30459-61	10/02/2019	1010A
10/09/2019	Water	K5180-01 <u>E</u>	Flash Point	Cool 4 deg C	PSEG01	N21	30474	10/02/2019	1010A
10/09/2019	Water	K5180-02 <u>E</u>	Flash Point	Cool 4 deg C	PSEG01	N21	40092	10/02/2019	1010A
10/15/2019	Water	K5185-01 <u>K</u>	Flash Point	Cool 4 deg C	AREC01	O32	WC-1	10/02/2019	1010A

Date/Time 10/3/19 13:00
 Received by: KS
 Relinquished by: JP

Date/Time 10/3/19 15:55
 Received by: JP
 Relinquished by: KS