



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB105470

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP78553
Turbidity Calibration std, 0NTU	WP78552
Turbidity Calibration std, 20NTU	WP78548
Turbidity Calibration std, 40NTU	WP78547
Turbidity Calibration std, 1NTU	WP78551
Turbidity Calibration std, 10NTU	WP78549
Turbidity Calibration std, 5NTU	WP78550
10 NTU Standard 500 ml	W2489

Intercept: -0.0611

Slope: 0.9814

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.117	10/03/2019	12:42
2	CAL2	1	1	0.992	10/03/2019	12:51
3	CAL3	5	1	4.860	10/03/2019	12:56
4	CAL4	10	1	9.650	10/03/2019	13:04
5	CAL5	20	1	19.200	10/03/2019	13:10
6	CAL6	40	1	39.400	10/03/2019	13:15



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB105470

ANALYST: AMANDEEP
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.200	9.437	10/03/2019	13:18
2	ICB		1	0.110	0.174	10/03/2019	13:22
3	CCV1	10	1	9.580	9.824	10/03/2019	13:26
4	CCB1		1	0.099	0.163	10/03/2019	13:30
5	LB105470BL		1	0.108	0.172	10/03/2019	13:33
6	K5187-01		1	19.800	20.238	10/03/2019	13:36
7	K5187-01DUP		1	19.600	20.034	10/03/2019	13:40
8	CCV2	10	1	9.560	9.803	10/03/2019	13:44
9	CCB2		1	0.107	0.171	10/03/2019	13:48

LB 105470
WORKLIST(Hardcopy Internal Chain)

WorkList Name : TUR100319 WorkList ID : 131618 Department : Wet-Chemistry Date : 10-03-2019 11:03:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/04/2019	Water	K5187-01	✓ Turbidity	Cool 4 deg C	LIRO01	M21	UDS-6-01-100219	10/02/2019	SM2130 B

Date/Time 10-03-19 11:40
 Received by: AK
 Relinquished by: SR

Date/Time 10-03-19 13:50
 Received by: SR
 Relinquished by: SR