

Analysis Method: SM4500 S E

ANALYST: ketankumar

Parameter: Sulfide

SUPERVISOR REVIEW BY: apatel

Run Number: LB105477

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
Sodium thiosulfate titrant 0.025N	WP77620
IODINE SOLUTION .025N 1L	W2508
Starch Solution, 4L	W2379

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	T1 (mL)	T2 Initial	T2 Final	T2 Diff. (mL)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	AnalDate	Anal Time
1	PB123628BL		1	50	50	5.00	0.00	4.90	4.90	0.10	0.00	0.00	10/03/2019	14:45
2	PB123628BS	25	1	50	50	5.00	0.00	2.30	2.30	2.70	2.60	20.80	10/03/2019	14:48
3	K5111-10		1	50	50	5.00	0.00	4.84	4.84	0.16	0.06	0.48	10/03/2019	14:52
4	K5111-11		1	50	50	5.00	0.00	4.74	4.74	0.26	0.16	1.28	10/03/2019	14:55

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume