


LB105483

7471B

INSTRUMENT ID : CV1

Linear



A= 0.0000e+000

B= 1.7427e-004

C= 1.8363e-002

Rho= 0.9990652

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.027	0.027	52	0.000	52				
0.2	0.200	0.243	0.043	1288	0.0 %	1288				
2.5	2.500	2.574	0.074	14662	0.0 %	14662				
5.0	5.000	4.680	-0.320	26747	0.0 %	26747				
7.5	7.500	7.703	0.203	44097	0.0 %	44097				
10.0	10.000	9.973	-0.027	57121	0.0 %	57121				

LB105483 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 03 Oct 2019 16:17:58					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	52	-	0.00007471B	PPB		03 Oct 2019 16:20:27	S	Std
0.2 - 1	50-2	1288	-	0.20007471B	PPB		03 Oct 2019 16:22:45	S	Std
2.5 - 1	52-5	14662	-	2.50007471B	PPB		03 Oct 2019 16:25:03	S	Std
5.0 - 1	55-0	26747	-	5.00007471B	PPB		03 Oct 2019 16:27:21	S	Std
7.5 - 1	57-5	44097	-	7.50007471B	PPB		03 Oct 2019 16:29:38	S	Std
10.0 - 1	516-0	57121	-	10.00007471B	PPB		03 Oct 2019 16:39:49	S	Std
ICV41 - 1	ICV41	23310	4.0807	-7471B	PPB		03 Oct 2019 16:43:11	U	SMPL
ICB41 - 1	ICB41	8	0.0198	-7471B	PPB		03 Oct 2019 16:45:27	U	SMPL
CCV60 - 1	CCV60	27835	4.8693	-7471B	PPB		03 Oct 2019 16:47:42	U	SMPL
CCB60 - 1	CCB60	11	0.0203	-7471B	PPB		03 Oct 2019 16:49:58	U	SMPL
CRA - 1	CRA	1267	0.2392	-7471B	PPB		03 Oct 2019 16:52:14	U	SMPL
HighStd - 1	HighStd	55683	9.7225	-7471B	PPB		03 Oct 2019 16:54:30	U	SMPL
ChkStd - 1	ChkStd	40166	7.0183	-7471B	PPB		03 Oct 2019 16:56:46	U	SMPL
PB123631BL - 1	PBS	16	0.0212	-7471B	PPB		03 Oct 2019 16:59:04	U	SMPL
PB123631BS - 1	LCSS	21181	3.7097	-7471B	PPB		03 Oct 2019 17:01:22	U	SMPL
K5140-01 - 1	HR-05-100219	3625	0.6501	-7471B	PPB		03 Oct 2019 17:03:40	U	SMPL
K5143-01 - 1	NB-08-100219	6147	1.0896	-7471B	PPB		03 Oct 2019 17:05:57	U	SMPL
K5155-01 - 1	STOCKPILE	2602	0.4718	-7471B	PPB		03 Oct 2019 17:08:15	U	SMPL
K5155-01DUP - 1	STOCKPILEDUP	2591	0.4699	-7471B	PPB		03 Oct 2019 17:10:32	U	SMPL
K5155-01MS - 1	STOCKPILEMS	23080	4.0406	-7471B	PPB		03 Oct 2019 17:12:49	U	SMPL
CCV61 - 1	CCV61	27915	4.8832	-7471B	PPB		03 Oct 2019 17:15:05	U	SMPL
CCB61 - 1	CCB61	9	0.0199	-7471B	PPB		03 Oct 2019 17:17:21	U	SMPL
K5155-01MSD - 1	STOCKPILEMSD	24232	4.2414	-7471B	PPB		03 Oct 2019 17:19:37	U	SMPL
K5158-01 - 1	ROLL-OFF	3299	0.5933	-7471B	PPB		03 Oct 2019 17:21:52	U	SMPL
K5165-01 - 1	NBP-2019	26525	4.6410	-7471B	PPB		03 Oct 2019 17:24:09	U	SMPL
K5168-01 - 1	20190763-SS-COMPOSITE-3	1197	0.2270	-7471B	PPB		03 Oct 2019 17:26:25	U	SMPL
K5168-02 - 1	20190764-SS-COMPOSITE-4	1368	0.2568	-7471B	PPB		03 Oct 2019 17:28:42	U	SMPL
K5168-03 - 1	20190769-SS-COMPOSITE-9	5450	0.9682	-7471B	PPB		03 Oct 2019 17:30:59	U	SMPL
K5168-04 - 1	2019073-SS-COMPOSITE-13	5591	0.9927	-7471B	PPB		03 Oct 2019 17:33:16	U	SMPL
K5168-05 - 1	2019071-SS-COMPOSITE-11	4472	0.7977	-7471B	PPB		03 Oct 2019 17:35:34	U	SMPL
K5168-06 - 1	2019070-SS-COMPOSITE-10	4843	0.8624	-7471B	PPB		03 Oct 2019 17:37:52	U	SMPL
K5176-01 - 1	40166	13973	2.4535	-7471B	PPB		03 Oct 2019 17:40:08	U	SMPL
CCV62 - 1	CCV62	27907	4.8819	-7471B	PPB		03 Oct 2019 17:42:25	U	SMPL
CCB62 - 1	CCB62	47	0.0266	-7471B	PPB		03 Oct 2019 17:44:41	U	SMPL
K5176-02 - 1	30440-3077	308	0.0720	-7471B	PPB		03 Oct 2019 17:46:57	U	SMPL
K5176-05 - 1	30464-67	22760	3.9849	-7471B	PPB		03 Oct 2019 17:49:13	U	SMPL
K5176-15 - 1	30459-61	11985	2.1070	-7471B	PPB		03 Oct 2019 17:51:29	U	SMPL
K5181-01 - 1	SOUTH-AVENUE-1-ABCD	877	0.1712	-7471B	PPB		03 Oct 2019 17:53:46	U	SMPL
K5181-02 - 1	SOUTH-AVENUE-2-ABCD	1165	0.2214	-7471B	PPB		03 Oct 2019 17:56:03	U	SMPL
K5181-03 - 1	SOUTH-AVENUE-3-ABCD	938	0.1818	-7471B	PPB		03 Oct 2019 17:58:20	U	SMPL
K5181-04 - 1	SOUTH-AVENUE-4-ABCD	2055	0.3765	-7471B	PPB		03 Oct 2019 18:00:37	U	SMPL
K5155-01LX5 - 1		561	0.1161	-7471B	PPB		03 Oct 2019 18:02:55	U	SMPL
K5155-01A - 1		23834	4.1720	-7471B	PPB		03 Oct 2019 18:05:13	U	SMPL
CCV63 - 1	CCV63	28016	4.9008	-7471B	PPB		03 Oct 2019 18:07:31	U	SMPL
CCB63 - 1	CCB63	22	0.0222	-7471B	PPB		03 Oct 2019 18:09:48	U	SMPL