



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB105492

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/04/2019	11:20
2	K5180-04		1	9	100.00	>212.0	>212.0	10/04/2019	11:50
3	K5180-04DUP		1	9	100.00	>212.0	>212.0	10/04/2019	12:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 105492

WorkList Name : fp10419 WorkList ID : 131662 Department : Wet-Chemistry Date : 10-04-2019 08:49:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/09/2019	Solid	K5180-04 B	Flash Point	Cool 4 deg C	PSEG01	N21	30444	10/02/2019	1010A

Date/Time 10/4/19 11:00
 Received by: JP
 Relinquished by: JP

Date/Time 10/4/19 12:20
 Received by: JP
 Relinquished by: VS