



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105493

Slope : 98.6

pH Meter ID : WC PH METER

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/04/2019	12:05
2	CAL2	1	Water	NA	NA	20.1	7.00	10/04/2019	12:06
3	CAL3	1	Water	NA	NA	20.1	10.03	10/04/2019	12:07
4	ICV	1	Water	NA	NA	20.2	7.01	10/04/2019	12:10
5	CCV1	1	Water	NA	NA	20.2	2.02	10/04/2019	12:11
6	K5218-01	1	Water	NA	NA	20.7	4.70	10/04/2019	12:15
7	K5218-01DUP	1	Water	NA	NA	20.8	4.71	10/04/2019	12:16
8	CCV2	1	Water	NA	NA	20.1	12.02	10/04/2019	12:19

VB105493

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5218 WorkList ID : 131675 Department : Wet-Chemistry Date : 10-04-2019 11:59:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/02/2019	Water	K5218-01-1	pH	Cool 4 deg C	STER04	O51	POOL-WATER	10/02/2019	SM 4500-H B

Date/Time 10-04-19 12:00
 Received by: J
 Relinquished by: C

Date/Time 10-04-19 13:00
 Received by: C
 Relinquished by: J