



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB105495

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP78622
Turbidity Calibration std, 0NTU	WP78621
Turbidity Calibration std, 20NTU	WP78617
Turbidity Calibration std, 80NTU	WP78623
Turbidity Calibration std, 40NTU	WP78616
Turbidity Calibration std, 200NTU	WP78624
Turbidity Calibration std, 1NTU	WP78620
Turbidity Calibration std, 10NTU	WP78618
Turbidity Calibration std, 5NTU	WP78619
10 NTU Standard 500 ml	W2489

Intercept: 0.0807

Slope: 0.9371

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.110	10/04/2019	12:55
2	CAL2	1	1	0.997	10/04/2019	13:01
3	CAL3	5	1	4.610	10/04/2019	13:06
4	CAL4	10	1	9.220	10/04/2019	13:11
5	CAL5	20	1	18.600	10/04/2019	13:16
6	CAL6	40	1	37.700	10/04/2019	13:20
7	CAL7	80	1	75.800	10/04/2019	13:25
8	CAL8	200	1	187.200	10/04/2019	13:29



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB105495

ANALYST: AMANDEEP
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.180	9.710	10/04/2019	13:32
2	ICB		1	0.107	0.028	10/04/2019	13:36
3	CCV1	10	1	9.260	9.795	10/04/2019	13:39
4	CCB1		1	0.098	0.018	10/04/2019	13:44
5	LB105495BL		1	0.102	0.023	10/04/2019	13:48
6	K5217-01		1	112.300	119.752	10/04/2019	13:52
7	K5217-02		1	38.400	40.891	10/04/2019	13:55
8	K5217-02DUP		1	37.000	39.397	10/04/2019	13:58
9	CCV2	10	1	9.270	9.806	10/04/2019	14:02
10	CCB2		1	0.111	0.032	10/04/2019	14:06

WORKLIST(Hardcopy Internal Chain)

LB10 5495

WorkList Name : tur100419

WorkList ID : 131679

Department : Wet-Chemistry

Date : 10-04-2019 12:37:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/05/2019	Water	K5217-01	Turbidity	Cool 4 deg C	LIRO01	M31	UDS-4-26-100319	10/03/2019	SM2130 B
10/05/2019	Water	K5217-02	Turbidity	Cool 4 deg C	LIRO01	M31	UDS-4-16A-100319	10/03/2019	SM2130 B

Date/Time 10-04-19 12:45

Received by: AK

Relinquished by: SP

Date/Time 10-04-19 14:10

Received by: SP

Relinquished by: AK