



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 10/8/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 14:01
In Date: 10/07/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:25
Out Date: 10/08/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB105525

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K5223-01	BEPH3	1	1.13	9.54	7.42	74.8	
K5223-02	BEPD6	2	1.12	9.91	7.25	69.7	
K5223-03	BEPD7	3	1.13	9.6	6.41	62.3	
K5223-04	BEPD8	4	1.12	9.77	7.14	69.6	
K5223-05	BEPD9	5	1.12	9.8	6.6	63.1	
K5223-06	BEPE0	6	1.13	9.88	7.29	70.4	
K5223-07	BEPE3	7	1.13	9.7	7.28	71.8	
K5223-08	BEPE4	8	1.14	9.52	7.56	76.6	
K5223-09	BEPE1	9	1.14	9.78	6.02	56.5	
K5223-10	BEPE2	10	1.12	9.66	7.07	69.7	
K5223-11	BEPE5	11	1.12	9.9	7.34	70.8	
K5223-12	BEPH4	12	1.14	9.55	7.38	74.2	
K5223-13	BEPH5	13	1.13	9.5	7.83	80.0	
K5223-14	BEPH6	14	1.14	9.73	7.76	77.1	
K5223-15	BEPH7	15	1.13	9.52	8.17	83.9	
K5223-16	BEPA4	16	1.13	9.68	6.23	59.6	
K5223-17	BEPA5	17	1.14	9.8	5.69	52.5	
K5223-18	BEPA6	18	1.13	9.6	6.32	61.3	
K5223-19	BEPA7	19	1.14	9.66	6.82	66.7	
K5223-20	BEPA8	20	1.15	9.7	6.06	57.4	
K5223-21	BEPA8MS	21	1.15	9.7	6.06	57.4	
K5223-22	BEPA8MSD	22	1.15	9.7	6.06	57.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$