



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB105533

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/07/2019	11:25
2	K5215-12		1	9	60.00	140.0	140.0	10/07/2019	11:50
3	K5215-12DUP		1	9	62.00	143.6	143.6	10/07/2019	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 105533

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp10719 WorkList ID : 131745 Department : Wet-Chemistry Date : 10-07-2019 09:10:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/11/2019	Solid	K5215-12	Flash Point	Cool 4 deg C	PSEG01	N23	30473	10/04/2019	1010A

Date/Time 10/07/19 11:00
Received by: KS
Relinquished by: JP

Date/Time 10/7/19 12:25
Received by: JP
Relinquished by: KS