



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 10/8/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 13:02
In Date: 10/07/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:35
Out Date: 10/08/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB105535

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K5102-01	MBEPA9	1	1.15	9.64	6.88	67.5	
K5102-02	MBEPB0	2	1.16	9.8	6.88	66.2	
K5102-03	MBEPB1	3	1.14	9.8	7.69	75.6	
K5102-04	MBEPB2	4	1.13	9.57	6.6	64.8	
K5102-05	MBEPB3	5	1.15	9.64	7.45	74.2	
K5102-06	MBEPC4	6	1.15	9.76	7.24	70.7	
K5102-07	MBEPC5	7	1.15	9.89	7.8	76.1	
K5102-08	MBEPK6	8	1.15	9.79	7.1	68.9	
K5102-09	MBEPK7	9	1.12	9.62	7.38	73.6	
K5102-10	MBEPK8	10	1.14	9.92	7.5	72.4	
K5102-11	MBEPK9	11	1.14	9.7	7.46	73.8	
K5102-12	MBEPL0	12	1.12	9.67	7.92	79.5	
K5102-13	MBEPL1	13	1.14	9.82	8.14	80.6	
K5102-14	MBEPB4	14	1.12	9.97	6.47	60.5	
K5102-15	MBEPB5	15	1.12	9.56	6.12	59.2	
K5102-16	MBEPB6	16	1.16	9.76	6.39	60.8	
K5102-17	MBEPB7	17	1.13	9.8	7.25	70.6	
K5102-18	MBEPB8	18	1.13	9.75	7.93	78.9	
K5102-19	MBEPB9	19	1.15	9.7	6.98	68.2	
K5102-20	MBEPB9D	20	1.15	9.7	6.98	68.2	
K5102-21	MBEPB9S	21	1.15	9.7	6.98	68.2	
K5102-22	MBEPC1	22	1.15	9.76	7.49	73.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$