



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 10/8/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 13:02
In Date: 10/07/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:35
Out Date: 10/08/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB105536

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K5235-01	BEPA9	1	1.15	9.64	6.88	67.5	
K5235-02	BEPB0	2	1.16	9.8	6.88	66.2	
K5235-03	BEPB1	3	1.14	9.8	7.69	75.6	
K5235-04	BEPB2	4	1.13	9.57	6.6	64.8	
K5235-05	BEPB3	5	1.15	9.64	7.45	74.2	
K5235-06	BEPC4	6	1.15	9.76	7.24	70.7	
K5235-07	BEPC5	7	1.15	9.89	7.8	76.1	
K5235-08	BEPK6	8	1.15	9.79	7.1	68.9	
K5235-09	BEPK7	9	1.12	9.62	7.38	73.6	
K5235-10	BEPK8	10	1.14	9.92	7.5	72.4	
K5235-11	BEPK9	11	1.14	9.7	7.46	73.8	
K5235-12	BEPL0	12	1.12	9.67	7.92	79.5	
K5235-13	BEPL1	13	1.14	9.82	8.14	80.6	
K5235-14	BEPB4	14	1.12	9.97	6.47	60.5	
K5235-15	BEPB5	15	1.12	9.56	6.12	59.2	
K5235-16	BEPB6	16	1.16	9.76	6.39	60.8	
K5235-17	BEPB7	17	1.13	9.8	7.25	70.6	
K5235-18	BEPB8	18	1.13	9.75	7.93	78.9	
K5235-19	BEPB9	19	1.15	9.7	6.98	68.2	
K5235-20	BEPB9MS	20	1.15	9.7	6.98	68.2	
K5235-21	BEPB9MSD	21	1.15	9.7	6.98	68.2	
K5235-22	BEPC1	22	1.15	9.76	7.49	73.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$