


LB105539

7470A

INSTRUMENT ID: C U1

Linear



A= 0.0000e+000

B= 2.2997e-004

C= 2.5866e-002

Rho= 0.9998868

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.040	0.040	62	0.000	62				
0.2	0.200	0.239	0.039	927	0.0 %	927				
2.5	2.500	2.469	-0.031	10624	0.0 %	10624				
5.0	5.000	4.900	-0.100	21197	0.0 %	21197				
7.5	7.500	7.486	-0.014	32440	0.0 %	32440				
10.0	10.000	10.065	0.065	43657	0.0 %	43657				

LB105539 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 07 Oct 2019 12:37:13				
Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type
0.0 - 1	50.0	62	-	0.0000	7470A	PPB	07 Oct 2019 12:50:04	S Std
0.2 - 1	50.2	927	-	0.2000	7470A	PPB	07 Oct 2019 12:52:13	S Std
2.5 - 1	50.5	10624	-	2.5000	7470A	PPB	07 Oct 2019 12:54:22	S Std
5.0 - 1	55.0	21197	-	5.0000	7470A	PPB	07 Oct 2019 12:56:30	S Std
7.5 - 1	57.5	32440	-	7.5000	7470A	PPB	07 Oct 2019 12:58:39	S Std
10.0 - 1	510.0	43657	-	10.0000	7470A	PPB	07 Oct 2019 13:00:46	S Std
ICV42 - 1	ICV42	16687	3.8633	-	7470A	PPB	07 Oct 2019 13:06:05	U SMPL
ICB42 - 1	ICB42	18	0.0300	-	7470A	PPB	07 Oct 2019 13:08:11	U SMPL
CCV64 - 1	CCV64	21428	4.9536	-	7470A	PPB	07 Oct 2019 13:10:18	U SMPL
CCB64 - 1	CCB64	0	0.0259	-	7470A	PPB	07 Oct 2019 13:12:25	U SMPL
CRA - 1	CRA	937	0.2413	-	7470A	PPB	07 Oct 2019 13:14:32	U SMPL
HighStd - 1	HighStd	42514	9.8026	-	7470A	PPB	07 Oct 2019 13:16:39	U SMPL
ChkStd - 1	ChkStd	29800	6.8788	-	7470A	PPB	07 Oct 2019 13:18:46	U SMPL
PB123710BL - 1	PBW	-14	0.0226	-	7470A	PPB	07 Oct 2019 13:20:55	U SMPL
PB123710BS - 1	LCSW	16305	3.7754	-	7470A	PPB	07 Oct 2019 13:23:03	U SMPL
K5193-05 - 1	20190879-COMPOSITE-K	21	0.0307	-	7470A	PPB	07 Oct 2019 13:25:11	U SMPL
K5193-06 - 1	20190880-COMPOSITE-L	27	0.0321	-	7470A	PPB	07 Oct 2019 13:27:19	U SMPL
K5193-07 - 1	20190881-COMPOSITE-M	51	0.0376	-	7470A	PPB	07 Oct 2019 13:29:28	U SMPL
K5193-08 - 1	20190882-COMPOSITE-N	38	0.0346	-	7470A	PPB	07 Oct 2019 13:31:36	U SMPL
K5203-01 - 1	72-12016	32	0.0332	-	7470A	PPB	07 Oct 2019 13:33:43	U SMPL
CCV65 - 1	CCV65	21152	4.8901	-	7470A	PPB	07 Oct 2019 13:35:49	U SMPL
CCB65 - 1	CCB65	23	0.0312	-	7470A	PPB	07 Oct 2019 13:37:56	U SMPL
K5205-03 - 1	RBR-200052-200029	16	0.0295	-	7470A	PPB	07 Oct 2019 13:40:03	U SMPL
K5205-06 - 1	RBR-200028-RT-3425	26	0.0318	-	7470A	PPB	07 Oct 2019 13:42:10	U SMPL
K5205-06DUP - 1	RBR-200028-RT-3425DUP	19	0.0302	-	7470A	PPB	07 Oct 2019 13:44:18	U SMPL
K5205-06MS - 1	RBR-200028-RT-3425MS	15842	3.6690	-	7470A	PPB	07 Oct 2019 13:46:25	U SMPL
K5205-06MSD - 1	RBR-200028-RT-3425MSD	15511	3.5929	-	7470A	PPB	07 Oct 2019 13:48:33	U SMPL
PB123711BL - 1	PBW	14	0.0291	-	7470A	PPB	07 Oct 2019 13:50:42	U SMPL
PB123711BS - 1	LCSW	16246	3.7619	-	7470A	PPB	07 Oct 2019 13:52:50	U SMPL
K5141-08 - 1	JC-02-100419	67	0.0413	-	7470A	PPB	07 Oct 2019 13:54:59	U SMPL
K5141-08DUP - 1	JC-02-100419DUP	101	0.0491	-	7470A	PPB	07 Oct 2019 13:57:07	U SMPL
K5141-08MS - 1	JC-02-100419MS	15998	3.7048	-	7470A	PPB	07 Oct 2019 13:59:15	U SMPL
CCV66 - 1	CCV66	21260	4.9149	-	7470A	PPB	07 Oct 2019 14:01:22	U SMPL
CCB66 - 1	CCB66	13	0.0289	-	7470A	PPB	07 Oct 2019 14:03:30	U SMPL
K5141-08MSD - 1	JC-02-100419MSD	15807	3.6609	-	7470A	PPB	07 Oct 2019 14:05:37	U SMPL
K5146-07 - 1	CL-01-100419	37	0.0344	-	7470A	PPB	07 Oct 2019 14:07:45	U SMPL
K5215-01 - 1	40302	73	0.0427	-	7470A	PPB	07 Oct 2019 14:09:52	U SMPL
K5230-01 - 1	TP-1	149	0.0601	-	7470A	PPB	07 Oct 2019 14:12:00	U SMPL
K5230-02 - 1	TP-2	103	0.0496	-	7470A	PPB	07 Oct 2019 14:15:07	U SMPL
PB123712BL - 1	PBW	54	0.0383	-	7470A	PPB	07 Oct 2019 14:17:14	U SMPL
PB123712BS - 1	LCSW	16387	3.7943	-	7470A	PPB	07 Oct 2019 14:19:22	U SMPL
K5154-21 - 1	MW-102-S	72	0.0424	-	7470A	PPB	07 Oct 2019 14:21:31	U SMPL
K5154-21DUP - 1	MW-102-SDUP	106	0.0502	-	7470A	PPB	07 Oct 2019 14:23:41	U SMPL
K5154-21MS - 1	MW-102-SMS	14258	3.3047	-	7470A	PPB	07 Oct 2019 14:25:48	U SMPL
CCV67 - 1	CCV67	21142	4.8878	-	7470A	PPB	07 Oct 2019 14:27:59	U SMPL
CCB67 - 1	CCB67	-12	0.0231	-	7470A	PPB	07 Oct 2019 14:30:05	U SMPL
K5154-21MSD - 1	MW-102-SMSD	15608	3.6152	-	7470A	PPB	07 Oct 2019 14:32:12	U SMPL
K5154-22 - 1	MW-102-I	82	0.0447	-	7470A	PPB	07 Oct 2019 14:34:20	U SMPL
K5154-23 - 1	MW-102-D	223	0.0771	-	7470A	PPB	07 Oct 2019 14:36:27	U SMPL
K5154-24 - 1	MW-101-S	126	0.0548	-	7470A	PPB	07 Oct 2019 14:38:35	U SMPL
K5154-25 - 1	MW-101-I	60	0.0397	-	7470A	PPB	07 Oct 2019 14:40:43	U SMPL
K5154-26 - 1	MW-101-D	144	0.0590	-	7470A	PPB	07 Oct 2019 14:42:51	U SMPL
K5213-09 - 1	W-1	30599	7.0626	-	7470A	PPB	07 Oct 2019 14:45:00	U SMPL
PB123713BL - 1	PBW	49	0.0371	-	7470A	PPB	07 Oct 2019 14:53:20	U SMPL
PB123713BS - 1	LCSW	16127	3.7345	-	7470A	PPB	07 Oct 2019 14:55:28	U SMPL
K5111-09 - 1	MDL-MDL-WATER-03-QT4-2019	164	0.0636	-	7470A	PPB	07 Oct 2019 15:06:27	U SMPL
CCV68 - 1	CCV68	21698	5.0156	-	7470A	PPB	07 Oct 2019 15:08:35	U SMPL
CCB68 - 1	CCB68	2	0.0263	-	7470A	PPB	07 Oct 2019 15:10:42	U SMPL
PB123710TB - 1	PB123623TB	55	0.0385	-	7470A	PPB	07 Oct 2019 15:12:49	U SMPL
PB123711TB - 1	PB123662TB	86	0.0456	-	7470A	PPB	07 Oct 2019 15:14:57	U SMPL
K5205-06LX5 - 1		1	0.0261	-	7470A	PPB	07 Oct 2019 15:17:04	U SMPL
K5205-06A - 1		15615	3.6168	-	7470A	PPB	07 Oct 2019 15:19:12	U SMPL
K5141-08LX5 - 1		10	0.0282	-	7470A	PPB	07 Oct 2019 15:21:20	U SMPL
K5141-08A - 1		16056	3.7182	-	7470A	PPB	07 Oct 2019 15:23:28	U SMPL
K5154-21LX5 - 1		2	0.0263	-	7470A	PPB	07 Oct 2019 15:25:36	U SMPL
K5154-21A - 1		15040	3.4845	-	7470A	PPB	07 Oct 2019 15:27:45	U SMPL
CCV69 - 1	CCV69	21586	4.9899	-	7470A	PPB	07 Oct 2019 15:29:53	U SMPL
CCB69 - 1	CCB69	-4	0.0249	-	7470A	PPB	07 Oct 2019 15:32:01	U SMPL