



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB105545

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/07/2019	17:00
2	CAL2	1	Water	NA	NA	20.1	6.99	10/07/2019	17:01
3	CAL3	1	Water	NA	NA	20.1	10.02	10/07/2019	17:02
4	ICV	1	Water	NA	NA	20.1	7.00	10/07/2019	17:05
5	CCV1	1	Water	NA	NA	20.1	2.02	10/07/2019	17:06
6	K5147-04	1	Solid	20.02	20	22.7	7.97	10/07/2019	17:07
7	K5152-05	1	Solid	20.04	20	22.9	8.50	10/07/2019	17:09
8	K5238-06	1	Solid	20.02	20	22.6	6.49	10/07/2019	17:10
9	K5238-06DUP	1	Solid	20.01	20	22.6	6.50	10/07/2019	17:11
10	CCV2	1	Water	NA	NA	20.2	12.02	10/07/2019	17:15

2-B 105545

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K5238 WorkList ID : 131773 Department : Wet-Chemistry Date : 10-07-2019 14:36:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/07/2019	Solid	K5147-04	Corrosivity	Cool 4 deg C	PSEG05	N23	OK-03-100719	10/07/2019	9045D
10/07/2019	Solid	K5152-05	Corrosivity	Cool 4 deg C	PSEG05	N23	SU-02-100719	10/07/2019	9045D
10/07/2019	Solid	K5238-06	Corrosivity	Cool 4 deg C	PSEG01	N23	TP-COMP	10/07/2019	9045D

Date/Time 10/07/19 15:00
Received by: Jo
Relinquished by: CS

Date/Time 10/07/19 17:00
Received by: AO
Relinquished by: JS