


LB105556

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.1079e-004

C= -3.6696e-002

Rho= 0.9999377

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.007	-0.007	141	0.000	141					—
0.05	0.050										
0.20	0.200	0.188	-0.012	1067	0.0 %	1067					
2.50	2.500	2.472	-0.028	11903	0.0 %	11903					
5.00	5.000	5.091	0.091	24325	0.0 %	24325					
7.50	7.500	7.480	-0.020	35659	0.0 %	35659					
10.0	10.000	9.976	-0.024	47501	0.0 %	47501					

%D

—

-6

-1

200

0

RS 10.08.19

LB105556
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	141	-		0 ISM02.4	PPB	10/8/2019 10:28	Std
	0.2 S0.2	1067	-		0.2 ISM02.4	PPB	10/8/2019 10:31	Std
	2.5 S2.5	11903	-		2.5 ISM02.4	PPB	10/8/2019 10:33	Std
	5 S5.0	24325	-		5 ISM02.4	PPB	10/8/2019 10:35	Std
	7.5 S7.5	35659	-		7.5 ISM02.4	PPB	10/8/2019 10:37	Std
	10 S10.0	47501	-		10 ISM02.4	PPB	10/8/2019 10:40	Std
ICV	ICV	18448	3.8519 -		ISM02.4	PPB	10/8/2019 10:49	SMPL
ICB	ICB	35	-0.0293 -		ISM02.4	PPB	10/8/2019 10:51	SMPL
CCV075	CCV075	24364	5.0989 -		ISM02.4	PPB	10/8/2019 10:53	SMPL
CCB075	CCB075	50	-0.0262 -		ISM02.4	PPB	10/8/2019 10:55	SMPL
PB123745BL	PBS745	73	-0.0213 -		ISM02.4	PPB	10/8/2019 10:58	SMPL
K5051-01	MJLMB4	817	0.1355 -		ISM02.4	PPB	10/8/2019 11:00	SMPL
K5051-02	MJLMB5	1279	0.2329 -		ISM02.4	PPB	10/8/2019 11:02	SMPL
K5051-03	MJLMB6	1026	0.1796 -		ISM02.4	PPB	10/8/2019 11:05	SMPL
K5051-04	MJLMB7	3291	0.657 -		ISM02.4	PPB	10/8/2019 11:07	SMPL
K5051-05	MJLMB9	1362	0.2504 -		ISM02.4	PPB	10/8/2019 11:09	SMPL
K5051-06	MJLMC0	1221	0.2207 -		ISM02.4	PPB	10/8/2019 11:11	SMPL
K5051-07	MJLMC1	2527	0.496 -		ISM02.4	PPB	10/8/2019 11:14	SMPL
K5051-08	MJLMC4	829	0.138 -		ISM02.4	PPB	10/8/2019 11:16	SMPL
K5051-09	MJLMC5	1078	0.1905 -		ISM02.4	PPB	10/8/2019 11:18	SMPL
K5051-10	MJLMC6	3348	0.669 -		ISM02.4	PPB	10/8/2019 11:20	SMPL
K5051-11	MJLMC9	1633	0.3075 -		ISM02.4	PPB	10/8/2019 11:23	SMPL
K5051-12	MJLMD0	1884	0.3604 -		ISM02.4	PPB	10/8/2019 11:25	SMPL
K5051-13	MJLMD1	2426	0.4747 -		ISM02.4	PPB	10/8/2019 11:27	SMPL
K5051-14	MJLMD2	2045	0.3944 -		ISM02.4	PPB	10/8/2019 11:30	SMPL
K5051-15	MJLMD3	787	0.1292 -		ISM02.4	PPB	10/8/2019 11:32	SMPL
K5051-16	MJLMD4	339	0.0348 -		ISM02.4	PPB	10/8/2019 11:34	SMPL
K5051-17	MJLMD5	1782	0.3389 -		ISM02.4	PPB	10/8/2019 11:36	SMPL
K5051-18	MJLMD6	806	0.1332 -		ISM02.4	PPB	10/8/2019 11:39	SMPL
K5051-19	MJLMD7	1359	0.2498 -		ISM02.4	PPB	10/8/2019 11:41	SMPL
CCV076	CCV076	24720	5.174 -		ISM02.4	PPB	10/8/2019 11:43	SMPL
CCB076	CCB076	19	-0.0327 -		ISM02.4	PPB	10/8/2019 11:46	SMPL
K5051-20	MJLME0	761	0.1237 -		ISM02.4	PPB	10/8/2019 11:48	SMPL
K5051-21	MJLME0D	751	0.1216 -		ISM02.4	PPB	10/8/2019 11:50	SMPL
K5051-22	MJLME0S	13288	2.7643 -		ISM02.4	PPB	10/8/2019 11:52	SMPL
PB123746BL	PBS746	31	-0.0302 -		ISM02.4	PPB	10/8/2019 11:55	SMPL
K5059-01	MDBAD5	14083	2.9318 -		ISM02.4	PPB	10/8/2019 11:57	SMPL
K5059-02	MDBAD8	1180	0.212 -		ISM02.4	PPB	10/8/2019 11:59	SMPL
K5059-03	MDBAE0	2274	0.4426 -		ISM02.4	PPB	10/8/2019 12:01	SMPL
K5059-04	MDBAE1	1160	0.2078 -		ISM02.4	PPB	10/8/2019 12:04	SMPL
K5059-05	MDBAE2	1427	0.2641 -		ISM02.4	PPB	10/8/2019 12:06	SMPL
K5059-06	MDBAE3	400	0.0476 -		ISM02.4	PPB	10/8/2019 12:08	SMPL
K5059-07	MDBAE5	8071	1.6646 -		ISM02.4	PPB	10/8/2019 12:11	SMPL
K5059-08	MDBAE6	1440	0.2668 -		ISM02.4	PPB	10/8/2019 12:13	SMPL
K5059-09	MDBAE9	2763	0.5457 -		ISM02.4	PPB	10/8/2019 12:15	SMPL
K5059-10	MDBAE9D	2780	0.5493 -		ISM02.4	PPB	10/8/2019 12:17	SMPL

LB105556
INSTRUMENT ID : CV1

K5059-11	MDBAE9S	16701	3.4837 -	ISM02.4	PPB	10/8/2019 12:20	SMPL
K5059-12	MDBAF0	1276	0.2323 -	ISM02.4	PPB	10/8/2019 12:22	SMPL
K5059-13	MDBAF1	887	0.1503 -	ISM02.4	PPB	10/8/2019 12:24	SMPL
K5059-14	MDBAF2	190	0.0034 -	ISM02.4	PPB	10/8/2019 12:27	SMPL
K5059-15	MDBAF6	1032	0.1808 -	ISM02.4	PPB	10/8/2019 12:29	SMPL
K5059-16	MDBAG5	703	0.1115 -	ISM02.4	PPB	10/8/2019 12:31	SMPL
CCV077	CCV077	24709	5.1717 -	ISM02.4	PPB	10/8/2019 12:33	SMPL
CCB077	CCB077	33	-0.0297 -	ISM02.4	PPB	10/8/2019 12:36	SMPL
K5059-17	MDBAG6	748	0.121 -	ISM02.4	PPB	10/8/2019 12:38	SMPL
K5059-18	MDBAG7	248	0.0156 -	ISM02.4	PPB	10/8/2019 12:40	SMPL
K5059-19	MDBAH0	1080	0.191 -	ISM02.4	PPB	10/8/2019 12:43	SMPL
K5059-20	MDBAH1	655	0.1014 -	ISM02.4	PPB	10/8/2019 12:45	SMPL
K5059-21	MDBAH2	507	0.0702 -	ISM02.4	PPB	10/8/2019 12:47	SMPL
K5059-22	MDBAH4	394	0.0464 -	ISM02.4	PPB	10/8/2019 12:49	SMPL
K5059-23	MDBAH7	1139	0.2034 -	ISM02.4	PPB	10/8/2019 12:52	SMPL
PB123747BL	PBS747	59	-0.0243 -	ISM02.4	PPB	10/8/2019 12:54	SMPL
K5063-01	MDBAD6	559936	HIGH -	ISM02.4	PPB	10/8/2019 12:56	
K5063-02	MDBAH8	1657	0.3126 -	ISM02.4	PPB	10/8/2019 12:58	SMPL
K5063-03	MDBAJ0	942	0.1619 -	ISM02.4	PPB	10/8/2019 13:00	SMPL
K5063-04	MDBAJ2	7578	1.5607 -	ISM02.4	PPB	10/8/2019 13:03	SMPL
K5063-05	MDBAJ3	527	0.0744 -	ISM02.4	PPB	10/8/2019 13:05	SMPL
K5063-06	MDBAJ5	2823	0.5584 -	ISM02.4	PPB	10/8/2019 13:07	SMPL
K5063-07	MDBAJ6	614	0.0927 -	ISM02.4	PPB	10/8/2019 13:10	SMPL
K5063-08	MDBAJ9	612	0.0923 -	ISM02.4	PPB	10/8/2019 13:12	SMPL
K5063-09	MDBAK2	27129	5.6818 -	ISM02.4	PPB	10/8/2019 13:14	SMPL
K5063-10	MDBAK2D	27358	5.73 -	ISM02.4	PPB	10/8/2019 13:16	SMPL
K5063-11	MDBAK2S	42964	9.0196 -	ISM02.4	PPB	10/8/2019 13:19	SMPL
K5063-12	MDBAK3	25775	5.3964 -	ISM02.4	PPB	10/8/2019 13:21	SMPL
CCV078	CCV078	25093	5.2526 -	ISM02.4	PPB	10/8/2019 13:23	SMPL
CCB078	CCB078	436	0.0552 -	ISM02.4	PPB	10/8/2019 13:25	SMPL
K5063-13	MDBAK4	13384	2.7845 -	ISM02.4	PPB	10/8/2019 13:28	SMPL
K5063-14	MDBAK5	9812	2.0316 -	ISM02.4	PPB	10/8/2019 13:30	SMPL
K5063-15	MDBAK6	43240	9.0778 -	ISM02.4	PPB	10/8/2019 13:32	SMPL
K5063-16	MDBAK7	2101	0.4062 -	ISM02.4	PPB	10/8/2019 13:35	SMPL
K5063-17	MDBAK9	661	0.1026 -	ISM02.4	PPB	10/8/2019 13:37	SMPL
K5063-18	MDBAL1	897	0.1524 -	ISM02.4	PPB	10/8/2019 13:39	SMPL
K5063-19	MDBAF3	191347	40.2969 -	ISM02.4	PPB	10/8/2019 13:41	SMPL
K5063-20	MDBAF4	2219	0.431 -	ISM02.4	PPB	10/8/2019 13:44	SMPL
K5063-21	MDBAF5	354	0.0379 -	ISM02.4	PPB	10/8/2019 13:46	SMPL
K5063-22	MDBAK8	94555	19.8944 -	ISM02.4	PPB	10/8/2019 13:50	SMPL
PB123748BL	PBS748	-24	-0.0418 -	ISM02.4	PPB	10/8/2019 13:52	SMPL
K5066-01	MDBAF7	19418	4.0564 -	ISM02.4	PPB	10/8/2019 13:55	SMPL
K5066-02	MDBAF8	4522	0.9165 -	ISM02.4	PPB	10/8/2019 13:57	SMPL
K5066-03	MDBAF9	5492	1.121 -	ISM02.4	PPB	10/8/2019 13:59	SMPL
K5066-04	MDBAF9D	5456	1.1134 -	ISM02.4	PPB	10/8/2019 14:01	SMPL
K5066-05	MDBAF9S	16424	3.4253 -	ISM02.4	PPB	10/8/2019 14:04	SMPL
K5066-06	MDBAG0	4387	0.888 -	ISM02.4	PPB	10/8/2019 14:06	SMPL

LB105556
INSTRUMENT ID : CV1

K5066-07	MDBAG2	417	0.0512	-	ISM02.4	PPB	10/8/2019 14:08	SMPL
K5066-08	MDBAG4	1372515	HIGH	-	ISM02.4	PPB	10/8/2019 14:11	
K5063-01DLX10		74977	15.7675	-	ISM02.4	PPB	10/8/2019 14:12	SMPL
CCV079	CCV079	23619	4.9419	-	ISM02.4	PPB	10/8/2019 14:16	SMPL
CCB079	CCB079	-1222	-0.2943	-	ISM02.4	PPB	10/8/2019 14:24	SMPL
K5063-01DL2X20		41945	8.8048	-	ISM02.4	PPB	10/8/2019 14:26	SMPL
K5063-19DLX10		19395	4.0515	-	ISM02.4	PPB	10/8/2019 14:28	SMPL
K5063-22DLX5		17247	3.5988	-	ISM02.4	PPB	10/8/2019 14:31	SMPL
K5066-08DLX10		132188	HIGH	-	ISM02.4	PPB	10/8/2019 14:33	
K5066-08DLX2X100		165527	34.8544	-	ISM02.4	PPB	10/8/2019 14:35	SMPL
K5066-08DL3X1000		17389	3.6287	-	ISM02.4	PPB	10/8/2019 14:43	SMPL
CCV080	CCV080	25260	5.2878	-	ISM02.4	PPB	10/8/2019 14:45	SMPL
CCB080	CCB080	-351	-0.1107	-	ISM02.4	PPB	10/8/2019 14:50	SMPL