



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105578

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/08/2019	17:00
2	CAL2	1	Water	NA	NA	20.1	7.01	10/08/2019	17:01
3	CAL3	1	Water	NA	NA	20.1	10.03	10/08/2019	17:02
4	ICV	1	Water	NA	NA	20.3	7.00	10/08/2019	17:05
5	CCV1	1	Water	NA	NA	20.1	2.02	10/08/2019	17:16
6	K5281-01	1	Solid	20.02	20	22.7	5.85	10/08/2019	17:17
7	K5281-01DUP	1	Solid	20.03	20	22.8	5.86	10/08/2019	17:18
8	CCV2	1	Water	NA	NA	20.3	12.02	10/08/2019	17:20

20105578

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5281 WorkList ID : 131840 Department : Wet-Chemistry Date : 10-08-2019 16:41:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/08/2019	Solid	K5281-01	pH	Cool 4 deg C	TULL02	P21	FK-PS-01-062	10/08/2019	9045D

Date/Time 10.08.19 15:45
Received by: JS
Relinquished by: CS

Date/Time 10/08/19 18:00
Received by: JS
Relinquished by: JS