



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 10/10/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 15:00
In Date: 10/09/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 102
Time OUT: 07:30
Out Date: 10/10/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB105584

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K5265-01	DBAR1	1	1.14	6.42	5.72	86.7	
K5265-02	DBAR1MS	2	1.14	6.42	5.72	86.7	
K5265-03	DBAR1MSD	3	1.14	6.42	5.72	86.7	
K5265-04	DBAR3	4	1.18	4.68	4.03	81.4	
K5265-05	DBAR4	5	1.15	9.03	7.44	79.8	
K5265-06	DBAR5	6	1.12	7.6	6.36	80.9	
K5265-07	DBAR6	7	1.12	5.34	4.43	78.4	
K5265-08	DBAR7	8	1.16	7.05	6.03	82.7	
K5265-09	DBAR8	9	1.17	4.92	4.27	82.7	
K5265-10	DBAR9	10	1.16	6.45	5.87	89.0	
K5265-11	DBAS0	11	1.11	5.96	5.7	94.6	
K5265-12	DBAS3	12	1.15	5.54	4.71	81.1	
K5265-13	DBAS4	13	1.12	6.33	5.37	81.6	
K5265-14	DBAS5	14	1.19	6.7	5.52	78.6	
K5265-15	DBAS6	15	1.13	5.36	5.08	93.4	
K5265-16	DBAS7	16	1.14	5.16	4.95	94.8	
K5265-17	DBAS9	17	1.18	9.8	8.01	79.2	
K5265-18	DBAT0	18	1.17	7.9	6.71	82.3	
K5265-19	DBAT1	19	1.1	6.45	5.61	84.3	
K5265-20	DBAT2	20	1.16	8.02	7.04	85.7	
K5265-21	DBAT3	21	1.16	9.88	8.98	89.7	
K5265-22	DBAT4	22	1.15	5.74	5.6	96.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$